410B DC Power Supply

Instruction Manual

P/N 294009 December 1965



WARRANTY

Notwithstanding any provision of any agreement the following warranty is exclusive:

The JOHN FLUKE MFG. CO., INC., warrants each instrument it manufactures to be free from defects in material and workmanship under normal use and service for the period of 1-year from date of purchase. This warranty extends only to the original purchaser. This warranty shall not apply to fuses, disposable batteries (rechargeable type batteries are warranted for 90-days), or any product or parts which have been subject to misuse, neglect, accident, or abnormal conditions of operations.

In the event of failure of a product covered by this warranty, John Fluke Mfg. Co., Inc., will repair and callbrate an instrument returned to an authorized Service Facility within 1 year of the original purchase; provided the warrantor's examination discloses to its satisfaction that the product was defective. The warrantor may, at its option, replace the product in lieu of repair. With regard to any instrument returned within 1 year of the original purchase, said repairs or replacement will be made without charge. If the failure has been caused by misuse, neglect, accident, or abnormal conditions of operations, repairs will be billed at a nominal cost. In such case, an estimate will be submitted before work is started, if requested.

THE FOREGOING WARRANTY IS IN LIEU OF ALL OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTY OF MERCHANTABILITY, FITNESS, OR ADEQUACY FOR ANY PARTICULAR PURPOSE OR USE. JOHN FLUKE MFG. CO., INC., SHALL NOT BE LIABLE FOR ANY SPECIAL, INCIDENTAL, OR CONSEQUENTIAL DAMAGES, WHETHER IN CONTRACT, TORT, OR OTHERWISE.

If any failure occurs, the following steps should be taken:

- 1. Notify the JOHN FLUKE MFG. CO., INC., or nearest Service facility, giving full details of the difficulty, and include the model number, type number, and serial number. On receipt of this information, service data, or shipping instructions will be forwarded to you.
- 2. On receipt of the shipping instructions, forward the instrument, transportation prepaid. Repairs will be made at the Service Facility and the instrument returned, transportation prepaid.

SHIPPING TO MANUFACTURER FOR REPAIR OR ADJUSTMENT

All shipments of JOHN FLUKE MFG. CO., INC., instruments should be made via United Parcel Service or "Best Way"* prepaid. The instrument should be shipped in the original packing carton; or if it is not available, use any suitable container that is rigid and of adequate size. If a substitute container is used, the instrument should be wrapped in paper and surrounded with at least four Inches of excelsior or similar shock-absorbing material.

CLAIM FOR DAMAGE IN SHIPMENT TO ORIGINAL PURCHASER

The instrument should be thoroughly inspected immediately upon original delivery to purchaser. All material in the container should be checked against the enclosed packing list. The manufacturer will not be responsible for shortages against the packing sheet unless notified immediately. If the instrument is damaged in any way, a claim should be filed with the carrier immediately. (To obtain a quotation to repair shipment damage, contact the nearest Fluke Technical Center.) Final claim and negotiations with the carrier must be completed by the customer.

The JOHN FLUKE MFG. CO., INC, will be happy to answer all applications or use questions, which will enhance your use of this instrument. Please address your requests or correspondence to: JOHN FLUKE MFG. CO., INC., P.O. BOX C9090, EVERETT, WASHINGTON 98206, ATTN: Sales Dept. For European Customers: Fluke (Holland) B.V., P.O. Box 5053, 5004 EB, Tilburg, The Netherlands.

*For European customers, Air Freight prepaid.

John Fluke Mfg. Co., Inc., P.O. Box C9090, Everett, Washington 98206

Rev. 6/81

Change/Errata Information

Issue No: 4 6/79

This change/errata contains information necessary to ensure the accuracy of the following manual. Enter the corrections in the manual if either one of the following conditions exist:

- 1. The instrument's pcb revision letter is equal to or higher than that which is indicated at the beginning of the change.
- 2. No revision letter is indicated at the beginning of the change/errata.

MANUAL

Title:

MODEL 4108 HIGH VOLTAGE DC POWER SUPPLY

Print Date:

DECEMBER 1, 1965

Rev and Date: _____

C/E PAGE EFFECTIVITY						
Page No.	Print Date					
1	11/77					
2	2/78					
3	6/79					

ERRATA #1

On page 2-2 paragraph 2-7e, change CAUTION note to read as follows:

CAUTION!

The sample string resistors in the 410B are subject to damage if the output voltage is reduced too rapidly. Pause approximately ½ second in each switch position when reducing the setting of the first decade switch (1000 volt increments). The second decade should be set to 300 or greater or the HIGH VOLTAGE switch set to STANDBY RESET before the first decade is set to 0 (zero).

CHANGE #1-3091

On page 5-10, remove Q204 from Ref Desig group Q203 and Q204 and add it to group Q205 and Q206 (use Code D). Change the tot qty of Q203 from 3 to 2. Change the tot qty of Q205 and Q206 from 2 to 3.


CHANGE #2-6647

On page 5-9 delete the entire C215 entry and add the following new entry: C215; Cap, optional or factory selected.

CHANGE #3-6720

On page 1-1, delete information under "INPUT POWER" and add: 100/115/230V ac, 50 to 500 Hz, approximatly 300 VA at full output.


On page 2-1, delete information under paragraph 2-4 and add: Two power transformers (T1) are available for this instrument, 100/115V ac or115/230V ac. On the rear panel of each instrument a decal indicates the operating voltage of the transformer installed. This operating voltage can be changed by altering the wiring configuration of the transformer. To change the operating voltage of the 115/230V ac transformer, refer to the decal located on the power transformer. Refer to the drawing below for changing the operating voltage of the 100/115V ac transformer. To gain access to the power transformer, remove the instrument's top cover and the Delay PCB Assembly.



On page 5-3, add after T1 description: (115/230V). Add another listing for T1 as follows: T1; Transformer, Aux. (100/115V); Stock No. -5600-350918; Mfr. -89536; Mfr. Part No. -5600-350918; Tot. Qty. -1.

On the schematic diagram, add, as a separate note, the following primary configuration of T1 for the 100/115V ac power transformer.

Transformer connections shown are for 100V ac operation. For 115V ac operation, move the wire at terminal 3 to terminal 4 or 7.



CHANGE #4-7609

On page 5-10, change the stock no, mfr. mfr part no, and tot qty for transistors Q211 and Q212 from: 4805-203489, 07910, CDQ 10656, REF to: 4819-168716, 07263, S19254, 2


CHANGE #5-8438

On page 5-7, add the following new entry: R155; Res, comp, 1.3k ± 5%, ½W; 4704-109157; 01121; EB1325; i.

CHANGE #6-415-1030

On page 5-11, add the following replacement for voltage reference 83A1 (V201). V201, 83A1 Substitute PCB, 451047, 89536, 451047.


On the schematic, remove and replace V201, R219 and R220 with the reference substitute shown below. Connect pin 1 to junction of R221 and base of Q206. Connect pin 2 to junction of DS203 and R141 (0V).



CHANGE #7 - 11910

Rev. -G, Amplifier PCB Assembly

On the Functional Schematic (High Voltage Power Supply) make the following changes:



CHANGE #8 - 12060

Rev. -H, Amplifier PCB Assembly

On page 5-10, make the following changes:

FROM: Q211/Transistor, NPN, Si/4819-168716/07263/519254/Ref TO: Q211/Transistor, NPN, Si/4805-203489/07910/CDQ10656/Ref

CHANGE #9 - 12329

Rev.-B, Front Panel Assembly (410B-406)

On page 5-6, make the following changes:

FROM: Handle, 6-7/16"/2404-101584/05704/805/2 TO: Handle/2404-494989/88245/1061-29/2

ERRATA #2

On page 1-1, under ELECTRICAL, add the following information:

Protection Class #1 (Relates solely to insulation or grounding properties further defined in IEC 348.)

TABLE OF CONTENTS

Section	Title	Pag
I	INTRODUCTION AND SPECIFICATIONS	1-1
·	1-1. Introduction	1-1 1-1 1-1
п	OPERATING INSTRUCTIONS	2-1
	2-1. Function of External Controls, Terminals, and Indicators. 2-3. Input Power. 2-5. Initial Operation. 2-7. General Purpose Use. 2-8. Use As A Calibrator. 2-10. Direct Calibration. 2-11. Precision Calibration. 2-12. Notes on Operation. 2-13. Meter Accuracy. 2-15. Overload Protection. 2-17. Polarity Switching. 2-19. Coaxial Cable Assembly. 2-22. Dage Type 486-1 Connector Assembly. 2-23. Amphenol MS 3106A Connector Assembly.	2-1 2-2 2-2 2-2 2-3 2-3 2-3 2-3 2-3 2-4 2-4 2-4 2-4
ın	THEORY OF OPERATION	3-1
	3-1. Introduction	3-1 3-1 3-1 3-2 3-2 3-2 3-2 3-2
IV	3-18. Accuracy	3-2 4-1
	4-1. Introduction	4-1

Rev. 1

TABLE OF CONTENTS, continued

Section	Title	Page
v	LIST OF REPLACEABLE PARTS	5-1
	5-1. Introduction 5-4. How To Obtain Parts 5-7. Abbreviations 5-8. Prefix Symbols 5-9. Quantity Symbols 5-10. Special Notes and Symbols 5-11. Use Code Effectivity	5-1 5-1 5-2 5-2 5-2 5-2 5-2 5-15
VI	NOT APPLICABLE	
VII	GENERAL INFORMATION	7-1
	7-1. List of Abbreviations	7-3 7-10 7-11
VIII	SCHEMATIC DIAGRAMS	8-1

LIST OF ILLUSTRATIONS

Figure	Title	Page
Frontispiece	Model 410B High Voltage DC Power Supply	iv
2-1.	Function of External Controls, Terminals and Indicators	2-1
2-2.	Typical Output Voltage Increment vs. Overcurrent Trip Setting	2-3
3-1.	Model 410B DC Power Supply - Block Diagram	3 – 1
4-1.	Troubleshooting	4-2
4-2.	Equipment Required for Troubleshooting and Calibration	4-3
5-1.	Final Assembly (Sheet 1 of 2)	5-4
5-1.	Final Assembly (Sheet 2 of 2)	5-5
5-2.	Front Panel Assembly	5-6
5-3.	Switch Circuit Board Assembly	5-8
5-4.	Amplifier Circuit Board Assembly	5-12
5-5.	Time Delay Circuit Board Assembly	5-13
5-6.	Rectifier Circuit Board Assembly	5-14



MODEL 410B HIGH VOLTAGE DC POWER SUPPLY

SECTION I

INTRODUCTION AND SPECIFICATIONS

INTRODUCTION 1-1.

GENERAL DESCRIPTION 1-2.

- The Fluke Model 410B High Voltage DC Power Supply is capable of providing an output of 0 to 10,000 vdc at 0 to 10 milliamperes. The output voltage is selected by four decade switches and a vernier potentiometer. The instrument may be operated with either positive or negative output terminal grounded, as selected by a front-panel switch. Overcurrent protection is provided to return the supply to standby operation in the event of excessive load current.
- The Model 410B is a hybrid design utilizing silicon transistors for the amplifier circuitry and a vacuum tube for the necessary high voltage series passing element. A time-delay relay is incorporated in the input circuitry, which prolongs the life of the high voltage passing tube. The reference element is a highly stable temperature compensated gas tube.
- The package design provides for adequate cooling at high ambient temperatures without the use of cooling fans. Rubber feet are provided for bench top operation to permit unrestricted air flow through the bottom cover air vents. The front panel is punched for mounting in a standard 19 inch rack. Side panels are tapped for Jonathan #130 quick disconnect chassis slides or other rack mounting supports.

RECEIVING INSPECTION 1-6.

This instrument has been thoroughly checked and tested before being shipped from the factory. Immediately after receiving the instrument, carefully inspect for damage which may have occurred in transit. If any damage is noted, follow the instructions outlined on the warranty page in the back of this manual.

SPECIFICATIONS

ELECTRICAL 1-9.

OUTPUT VOLTAGE: 0 to ±10,000 VDC.

OUTPUT CURRENT: 0 to 10 milliamperes.

OUTPUT POLARITY: + or - grounded via front panel switch.

LINE REGULATION: 0.001% or 2 mv (whichever is greater) for 10% line change from nominal.

LOAD REGULATION: 0.001% or 5 mv (whichever is greater) for full load change.

STABILITY: ±0.005% per hour; ±0.02% per day after warmup.

RESOLUTION: 5 millivolts.

RIPPLE: Less than 1 mv RMS; less than 5 mv peakto-peak.

VOLTAGE CALIBRATION:

0 to 9000V in 9 steps of 1000V

0 to 900V in 9 steps of 100V

90V in 9 steps of 10V 0 to 9V in 9 steps of 1V 0 to

0 to 1.2V vernier

CALIBRATION ACCURACY: ±0.25% or 250 my (whichever is greater) with vernier at zero.

RESETABILITY: ±0.05% or 50 mv (whichever is great-

RECOVERY TIME: Within 50 microseconds.

WARMUP TIME: 30 minutes.

OVERCURRENT TRIP: Set to latch off at 12 ma load current. Internally adjustable from 5 to 15 ma.

METER: 10,000-0-10,000 vdc (±3%).

OUTPUT CONNECTORS: MS3102A-18-16S front and rear (one mating connector supplied).

INPUT POWER: 115/230 VAC ±10%, 50 - 60 Hz, approximately 300 VA at full output. Operation at 400 Hz available upon request.

1-10. ME CHANICAL

HUMIDITY: 0 to 80%.

OPERATING TEMPERATURE RANGE: 0°C to 50°C.

STORAGE TEMPERATURE RANGE: -20°C to +70°C.

ALTITUDE, OPERATING: 0 to 10,000 ft.

ALTITUDE, NON-OPERATING: 0 to 50,000 ft.

VIBRATION: Meets MIL-T-945A.

SHOCK: Meets MIL-E-4970A (20 g's, 11 milliseconds in three principal axis).

TEMPERATURE COEFFICIENT OF OUTPUT: Less than 20 ppm per °C from +10°C to +40°C.

SIZE: 19" wide x 8-3/4" high x 15" behind panel (rack mount with resilient feet for bench use).

WEIGHT: Approximately 59 pounds.




static awareness

A Message From

John Fluke Mfg. Co., Inc.






Some semiconductors and custom IC's can be damaged by electrostatic discharge during handling. This notice explains how you can minimize the chances of destroying such devices by:


- 1. Knowing that there is a problem.
- 2. Learning the guidelines for handling them.
- 3. Using the procedures, and packaging and bench techniques that are recommended.

The Static Sensitive (S.S.) devices are identified in the Fluke technical manual parts list with the symbol


The following practices should be followed to minimize damage to S.S. devices.




1. MINIMIZE HANDLING




2. KEEP PARTS IN ORIGINAL CONTAINERS UNTIL READY FOR USE.




3. DISCHARGE PERSONAL STATIC BEFORE HANDLING DEVICES. USE A HIGH RESISTANCE GROUNDING WRIST STRAP.




4. HANDLE S.S. DEVICES BY THE BODY



5. USE STATIC SHIELDING CONTAINERS FOR HANDLING AND TRANSPORT




6. DO NOT SLIDE S.S. DEVICES OVER ANY SURFACE




7. AVOID PLASTIC, VINYL AND STYROFOAM® IN WORK AREA

PORTIONS REPRINTED
WITH PERMISSION FROM TEKTRONIX, INC.
AND GENERAL DYNAMICS, POMONA DIV.



8. WHEN REMOVING PLUG-IN ASSEMBLIES, HANDLE ONLY BY NON-CONDUCTIVE EDGES AND NEVER TOUCH OPEN EDGE CONNECTOR EXCEPT AT STATIC-FREE WORK STATION. PLACING SHORTING STRIPS ON EDGE CONNECTOR HELPS TO PROTECT INSTALLED SS DEVICES.



- 9. HANDLE S.S. DEVICES ONLY AT A STATIC-FREE WORK STATION
- 10. ONLY ANTI-STATIC TYPE SOLDER-SUCKERS SHOULD BE USED.
- 11. ONLY GROUNDED TIP SOLDERING IRONS SHOULD BE USED.

A complete line of static shielding bags and accessories is available from Fluke Parts Department, Telephone 800-526-4731 or write to:

JOHN FLUKE MFG. CO., INC. PARTS DEPT. M/S 86 9028 EVERGREEN WAY EVERETT, WA 98204

SECTION II

OPERATING INSTRUCTIONS

2-1. <u>FUNCTION OF EXTERNAL CONTROLS, TERMINALS, AND INDICATORS</u>

2-2. The function of external controls, terminals, and indicators on the 410B is given in Figure 2-1.

2-3. INPUT POWER

2-4. Input power to the 410B is 115 vac or 230 vac ($\pm 10\%$), at a frequency of 50 to 60 Hz. Instruments are

supplied for operation from 400 Hz when requested. The 410B is equipped with dual primary windings on transformer T1, which are usually connected in parallel for operation from 115 vac. When requested, instruments are wired with the primary windings connected in series for operation from 230 vac. If it becomes desirable to change from one operating voltage to the other, remove the top cover of the instrument, and change the jumper wires and the fuse as shown on the transformer decal.

CONTROL TERMINAL, OR INDICATOR	LOCATION	REFERENCE DESIGNATION	
POWER switch	Front panel	S1	Applies line power to the control circuit and to the auxiliary transformer.
POWER lamp	Front panel	DS1	Indicates the application of line power to the instrument circuitry.
Fuse	Front panel	F1	Protects the instrument against damage due to overload.
STANDBY- RESET lamp	Front panel	DS2	This lamp illuminates when the time delay cycle is completed when the HIGH VOLTAGE switch is set to STANDBY-RESET.
HIGH VOLTAGE switch	Front panel	S2	Energizes and de-energizes the primary of transformer T2.
HIGH VOLTAGE ON lamp	Front panel	DS3	This lamp illuminates when the HIGH VOLTAGE switch is set to ON (after the time-delay cycle is completed), indicating that the high voltage is available at the output connectors.
POLARITY switch	Front panel	S7	Used to select either a positive or negative output with respect to chassis ground. This switch must be pulled out to turn, which interrupts ac input power to the instrument.
OUTPUT VOLTAGE controls	Front panel	S3, S4, S5, S6, & R140	Switches S3, S4, S5, & S6 select the output voltage in steps of 1000V, 100V, 10V, and 1V, respectively. R140 provides a continuous output of 0 to 1.2 volts.
Voltmeter	Front panel	M1	Indicates the output voltage from -10.0 to +10.0 KV. Accuracy is approximately 3% of end scale.
OUTPUT connector	Front and back panels	J1 and J2	Provided for connecting the load circuit to the 410B.

Figure 2-1. FUNCTION OF EXTERNAL CONTROLS, TERMINALS, AND INDICATORS

2-5, INITIAL OPERATION

- 2-6. The following procedure is recommended when turning on the Model 410B for the first time after shipping or a long period of idleness. This procedure will minimize the possibility of accidental damage to the unit.
- a. The 410B is shipped with the series passing tube removed from the socket and packed in a carton, which is fastened to the top cover. Install this tube in its socket and connect the plate wire to the plate cap before applying power to the instrument.

b. Connect the line plug to a 115 vac power source

(or 230 vac if so wired).

WARNING

The round pin on the polarized three-prong plug connects the instrument case to power system ground. If a three-to-two pin adapter is used to connect to a two-contact outlet, connect the short lead on the adapter to a good ground.

Always cover the unused output connector with the cover provided to prevent accidental contact with the high voltage, and dirt accumulation on the connector.

- c. Set the HIGH VOLTAGE switch to STANDBY-RESET.
- d. Set the POWER switch to ON. The POWER lamp will illuminate. After approximately 30 seconds, the time-delay relay will close and the STANDBY-RESET lamp will illuminate.

CAUTION

The 410B is designed for convection cooling without the use of fans. For bench top operation, the rubber feet must be installed or other steps taken to ensure adequate spacing for unrestricted air flow through the vents in the bottom cover.

e. Set the second OUTPUT VOLTAGE dial to 500. f. After the STANDBY-RESET lamp illuminates, set the HIGH VOLTAGE switch to ON. Carefully observe if the HIGH VOLTAGE lamp illuminates and if the output voltage rises to within 3% of 500 volts as indicated by the panel meter.

CAUTION

If the output voltage exceeds approximately 500 volts, immediately set the HIGH VOLTAGE switch to STANDBY-RESET.

- g. If the output voltage is 500 volts, the supply may be operated as in paragraph 2-7 or 2-8. If the output is not 500 volts, perform steps h. through j.
- h. Set the HIGH VOLTAGE switch to STANDBY-
- i. Locate and correct the source of trouble. Refer to Section IV.
- j. Repeat the initial operating procedure.

2-7. GENERAL PURPOSE USE

- a. Set the HIGH VOLTAGE switch to STANDBY-RESET.
- b. Set the POWER switch to ON. The POWER lamp will illuminate. After approximately 30 seconds, the time-delay relay will close and the STANDBY-RESET lamp will illuminate.
 - c. Set the POLARITY switch to the desired polarity.
- d. Connect the load to the power supply. Be sure the connection is firm.

WARNING

This instrument can produce lethal voltage. Always set the HIGH VOLTAGE switch to STANDBY-RESET and wait until the output decays to zero before connecting or disconnecting the load.

e. Set the OUTPUT VOLTAGE controls to the desired output. Check the external circuit for conflicts in grounding before applying power to the load.

CAUTION

The sampling string resistors in the 410B are subject to damage if the output voltage is reduced too rapidly. Pause approximately 1/2 second in each switch position when reducing the setting of the first decade switch (1000 volt increments).

f. After the STANDBY-RESET lamp illuminates, set the HIGH VOLTAGE switch to ON. The STANDBY-RESET lamp will extinguish and the HIGH VOLTAGE ON lamp will illuminate. The meter will indicate the approximate output voltage.

NOTE

If the overcurrent trip level has been reduced to less than 8 ma, it may be necessary to set the HIGH VOLTAGE switch to ON with a reduced output voltage to prevent actuating the overcurrent trip. In this case, step e., above, would follow step f. Figure 2-2 is a graph of the approximate maximum output voltage increment vs. the overcurrent trip setting. For example, if the overcurrent trip point has been reduced to 6 ma, the output voltage must be set to less than 3000 volts (approximately) to prevent actuating the overcurrent trip when the HIGH VOLT-AGE switch is set to ON. Figure 2-2 is typical, and is intended to show only the general characteristic.

g. To remove the high voltage from the output connectors, set the HIGH VOLTAGE switch to STANDBY-RESET.

2-8. USE AS A CALIBRATOR

2-9. The 410B may be used for direct calibration of dc instruments to an accuracy of better than 0. 25% or 250 millivolts at any voltage up to 10,000 volts. The




Figure 2-2. TYPICAL OUTPUT VOLTAGE INCREMENT VS. OVERCURRENT TRIP SETTING

four decade switches provide discrete steps with a resolution of one volt per step. The fifth dial provides a vernier with resolution of better than 5 millivolts for calibration between the cardinal points. When used with a Fluke differential voltmeter and voltage divider, the 410B is capable of calibrating dc instruments from 0 to 10,000 volts with an accuracy of 0.02% to 0.06%, depending upon the accuracy of the voltage divider and differential voltmeter used.

2-10. DIRECT CALIBRATION

- a. Allow the power supply sufficient time to warm up to stable operating temperature, usually not less than one half hour. If possible, set the OUTPUT VOLTAGE dials for the maximum required output during warmup in order to stabilize the operating temperature of all of the sample-string resistors which will be used during calibration. When used in this manner the power supply will provide an output significantly better in stability and accuracy than is specified.
- b. Set the POLARITY switch to the desired polarity.c. Set the HIGH VOLTAGE switch to STANDBY-
- RESET.
- d. Connect the instrument being calibrated to the OUTPUT connector.
- e. Set the OUTPUT VOLTAGE dials to the initial calibration point.
- f. Set the HIGH VOLTAGE switch to ON.
- g. Successively set the OUTPUT VOLTAGE dials to the desired calibration points.

2-11. PRECISION CALIBRATION

- a. Warm up supply as in paragraph 2-10a.
- b. Set the POLARITY switch to the desired polarity.
- c. Set the HIGH VOLTAGE switch to STANDBY-RESET.
- d. Connect the instrument being calibrated to the OUTPUT connector. Also connect a voltage divider, such as a Fiuke Modei 80B-10 to the OUTPUT connector.
- e. Connect a Fluke DC differential voltmeter to the output of the voltage divider. For calibration below 500 volts (1000 volts with the newer Fluke differential voltmeters) the voltage divider may be omitted.
- f. Set the differential voltmeter to differentially measure the voltage at the desired calibration point. When the voltage divider is used, consider the division ratio when setting the dials of the voltmeter.
 - g. Set the HIGH VOLTAGE switch to ON.
- h. Null the differential voltmeter by adjusting the power supply OUTPUT VOLTAGE dials. The accuracy of calibration is from 0.02% to 0.06%, depending on the accuracy of the differential voltmeter and voltage divider used.
- i. Repeat steps f. and h. for as many calibration points as desired.

2-12. NOTES ON OPERATION

2-13. METER ACCURACY

2-14. The meter in the 410B has a basic accuracy of $\pm 2\%$. Since multiplying and shunting resistors are used, the overall accuracy is approximately $\pm 3\%$ of end scale. However, the accuracy of the supply is $\pm 0.25\%$ of the output voltage. For example, if a voltage of 5000 volts is selected, the output voltage will be 5000 (± 12.5) volts ($\pm 0.25\%$ of $5000 = \pm 12.5$). However, the panel meter may indicate 5000 (± 300) volts ($\pm 3\%$ of $10,000 = \pm 300$) and still be within specified accuracy. Consequently, the front panel controls should be relied upon to indicate the magnitude of the output voltage.

2-15. OVERLOAD PROTECTION

2-16. The 410B is protected from damage due to overload by an overcurrent trip circuit. This circuit will return the supply to standby operation if the output current exceeds the pre-set level. This level is internally adjustable, and is usually set for 12 milliamperes,

but may be adjusted to lower levels, down to approximately 5 milliamperes. If tripped, the supply may be returned to operation by setting the HIGH VOLTAGE switch to STANDBY-RESET, waiting 30 seconds for the time-delay relay to close, and then setting the HIGH VOLTAGE switch to ON.

2-17. POLARITY SWITCHING

2-18. The output polarity of the 410B may be changed by pulling out the POLARITY switch until it reaches the stop, rotating the switch to the desired polarity, and pushing the switch all the way back to the stop. An interlock (S9) actuated by the switch shaft disables the instrument when the switch is pulled out. In order to return the supply to operation, the HIGH VOLTAGE switch must be set to STANDBY-RESET. After 30 seconds, when the STANDBY-RESET lamp illuminates, set the HIGH VOLTAGE switch to ON.

2-19. COAXIAL CABLE ASSEMBLY

2-20. Coaxial cables used to connect the instrument to a load are not provided by John Fluke Mfg. Co., Inc. Procurement of cables and the assembly of mating cable connectors must be carefully undertaken to insure safety of personnel.


WARNING!

Extreme care should be exercised when utilizing high voltage conductors and connectors. Improper assembly of cable connectors or careless use of high voltage equipment may expose personnel to voltage of lethal magnitude.


2-21. General instructions are provided herein for the assembly of two types of cable connectors issued as an accessory item with John Fluke instruments. Additional information on the assembly of these connectors, or on other types of connectors considered for use with the instrument, may be obtained from the cable or connector vending agency.

2-22. DAGE TYPE 486-1 CONNECTOR ASSEMBLY


a. Cut cable off square on end. Slide outer sleeve over cable and slide back out of way. Trim cable to dim. shown being careful not to nick center conductor or braid. Tin dip center conductor. Remove all excess solder.



b. Rotate dielectric to slightly flare braid.




c. Push cable into nut sub-assembly until the center conductor is visible thru hole in contact. Soft solder center conductor thru hole in contact. Remove excess solder. Fold braid down over back end tube of nut. Trim excess braid if necessary.




THE PERSON NAMED IN COLUMN 1

d. Slide outer sleeves over braid until it is flush against back of nut. Crimp as close to nut as possible using proper crimping tool.



e. Insert finished cable assy, into body either jack or plug and tighten by rotating body. Do not rotate nut sub-assembly,




2-23. AMPHENOL MS 3106A CONNECTOR ASSEMBLY

a. Strip cable jacket, braid, and dielectric to dimensions shown. All cuts are to be sharp and square. Do not nick braid, dielectric, and center conductor. Tinning of center conductor is not necessary if contact is to be crimped. For solder method, tin center conductor avoiding excessive heat.



b. Disassemble connector by removing the back shell and retainer ring.



WI	TH PIN INSERT (Male)	WIT	'H SOCKET INSERT (Female)
A	Back Shell	A	Back Shell
В	Retainer Ring	В	Retainer Ring
C	Pin Rear Insert	F	Socket Rear Insert
D	Pin Contact	G	Socket Contact
E	Pin Front Insert	Н	Socket Front Insert
J	Front Shell	J	Front Shell
K	Coupling Ring	K	Coupling Ring

- c. Insert cable thru clamp MS 3057A and rear components of connector. Place the cable center conductor into the socket contact.
- d. Soft solder contact to cable center conductor. Do not get any solder on outside surfaces of contact. Avoid excessive heat to prevent swelling or dielectric.

Contact must butt against cable dielectric



- e. Provide a connection from the cable shield to the securing bolt of the cable clamp. The connecting copper wire should be approximately 5 inches long and size 14 or larger in diameter. A terminal lug should be used to connect the wire to the clamp. The wire should be carefully soldered to the shield, taking care not to damage the insulation of the cable. Loose ends of the shield should be clipped off to prevent shorting the center conductor or contact to the shield.
- f. Assemble the cable connector by placing the contact into the socket front insert, then coupling the back shell to the front shell in the reverse of the disassembly procedure.

SECTION III

THEORY OF OPERATION

3-1. INTRODUCTION

3-2. This section of the manual describes the operation of the Model 410B. Reference is made to the functional schematic following Section V. This schematic is intended to aid in understanding the theory of operation, and in troubleshooting. A block diagram of the 410B is given in Figure 3-1. Operation of the 410B is discussed in the following paragraphs.

3-3. OVERALL OPERATION

3-4. The output voltage is controlled by the series passing tube V202, which is controlled by the feedback amplifier consisting of Q208 through Q214. Transistors Q213 and Q214 are connected as a differential amplifier, in which the output is proportional to the difference between the two inputs. The input of Q214 is connected to the summation point, which is the junction of the reference resistor R106, and the voltage control resistors R107 through R140. The feedback amplifier controls the output voltage so that the summation point is essentially at the same potential as the positive output bus. Any voltage difference between these two points is

amplified by the differential amplifier, and further amplified by Q212 through Q208, and applied to the grid of the series passing tube V202. The passing tube then increases or decreases its conductance until the voltage at the summation point is equal to the voltage of the positive bus. Tube V201 is a highly stable voltage reference which maintains a constant voltage of approximately 83 volts. Since the summation point is always held near zero volts, a constant current of 2 milliamperes flows through reference resistors R104 through R106, and also through voltage control resistors R107 through R140. The output voltage is equal to the IR drop across the voltage control resistors, and may thus be precisely controlled by varying the resistance of the voltage control string.

3-5. CIRCUIT DESCRIPTIONS

3-6. CONTROL RELAYS

3-7. A full-wave rectifier is formed by CR301 through CR304, which furnishes dc voltage to the control relays K301 and K302. Approximately 30 seconds after power is applied to the instrument, time-delay relay K301




Figure 3-1. MODEL 410B DC POWER SUPPLY - BLOCK DIAGRAM

closes if S2 is set to the STANDBY-RESET position. This energizes K302, which opens K302A, removing K301 from the circuit, and closes K302B and K302C. When switch S2 is set to the HIGH VOLTAGE ON position after K302C is closed, the primary of transformer T2 is energized and transformer voltage is applied to CR401 and CR402.

3-8. OVERCURRENT PROTECTION

3-9. Relay K303 provides overcurrent protection. If the output current exceeds approximately 12 milliamperes, K303 is actuated, which closes K303A. This de-energizes K302, which closes K302A, and opens K302B and K302C, extinguishing the HIGH VOLTAGE ON lamp. The primary circuit of T2 is thus opened by K302C, and high voltage is removed from the output terminals. Before high voltage can be reapplied to the output terminals, switch S2 must be set to the STANDBY-RESET position. After the time-delay cycle is complete, voltage will be available at the output connector when the HIGH VOLTAGE switch, S2, is set to ON.

3-10. MAIN OUTPUT VOLTAGE

3-11. The main output voltage is developed by CR401, CR402, C3, C4, and R402 through R411. This is a full-wave, voltage-doubler circuit, which applies rectified, filtered, dc voltage to the plate of V202. The taps on the primary of T2 provide approximate control of the unregulated dc voltage, which minimizes power dissipation in V202 at low output voltage.

3-12. FEEDBACK AMPLIFIER

3-13. The feedback amplifier consists of essentially five stages composed of seven silicon transistors. Transistors Q213 and Q214 are connected as a differential amplifier. The output is taken from the collector load of Q214 and is applied to the common-emitter amplifier Q212. Output from the collector of Q212 is ac coupled back to the base of Q213 as negative feedback for loop stabilization. The output of Q212 is applied to the input of the compound-connected, emitter-follower pair of Q210 and Q211. This stage provides current gain and impedance matching between the high collector impedance of Q212 and the low emitter impedance of Q209. Transistors Q209 and Q208 are common base stages which provide the voltage gain and large voltage swing required by the grid of V202. C215 and R230 in the emitter of Q209 form a lead network for loop compensation.

3-14. AUXILIARY SUPPLIES

3-15. AC voltage for both the positive and negative auxiliary supplies is obtained from one transformer winding. One terminal of this winding is connected to the positive output of the 410B, which is the zero volt reference. The other terminal of the winding is con-

nected to the junction of two half-wave rectifiers and filters. The positive auxiliary voltage is obtained from CR201, CR202, and C202. Auxiliary voltages of +95 volts, +20 volts, and +10 volts are obtained from CR205, CR206, and CR207, respectively, which are connected across the positive auxiliary voltage. The negative auxiliary voltage is obtained from CR203, CR204, and C203. An auxiliary voltage of -110 volts is obtained from CR208. This arrangement of the +95V, +20V, +10V, and -110V auxiliary supplies provides shunt regulation of these supplies.

3-16. REFERENCE VOLTAGE

3-17. Reference voltage for the instrument is obtained from V201, a gas-filled reference tube. Resistor R219 has a temperature coefficient of +0.45%/° C, which provides temperature compensation for V201. Transistors Q205 and Q206 are used as a differential amplifier. Transistor Q203 is operated as an emitter-follower voltage regulator, which provides 100 volts for the collector of Q205 and the emitter of Q204. The base of Q205 samples the output voltage of the +125 volt auxiliary supply, through the voltage divider R213 and R214. The output from Q206 represents the amplified difference between the reference voltage and the voltage at the base of Q205. Transistor Q204 provides further amplification of this voltage difference for application to series regulators Q202 and Q201. Neon lamps DS201 and DS202 are safety devices to prevent possible burn-out of Q201 and Q202 if the output of the instrument becomes shortcircuited.

3-18. ACCURACY

3-19. The main sampling string resistors in the 410B are accurate to within $\pm 0.1\%$. However, the accuracy of the 410B is specified as $\pm 0.25\%$, because the calibration accuracy also depends upon the stability of the reference voltage. The voltage of reference tube V201 (83A1) changes slightly due to aging. The accuracy of the supply will remain within $\pm 0.25\%$ for at least 30 days. The calibration accuracy may be maintained at better than 0.25% if the instrument is recalibrated more often than the usual calibration period of 30 days.

3-20. All calibrated power supplies have an accuracy limit (floor) as the output voltage approaches zero. This floor is caused by zero shift in the error amplifier, contact resistance in the sampling string circuit, and the accuracy of the sampling string resistors used for the least significant digits. The accuracy floor may be reduced by using more expensive components and additional circuitry. However, this results in greater initial cost and longer calibration time. Thus, very low accuracy floors are usually found only in precision calibrators. The 410B has an accuracy of $\pm 0.25\%$ or 250 millivolts, whichever is greater, when the first four voltage controls are used. Thus, the $\pm 0.25\%$ accuracy applies from 10,000 volts to 100 volts.

SECTION IV

MAINTENANCE

4-1. INTRODUCTION

4-2. Maintenance of the Model 410B DC Power Supply should consist primarily of occasional cleaning, tube replacement, and calibration. Preventive maintenance is discussed in paragraph 4-3. A discussion of trouble-shooting and a troubleshooting chart are presented in paragraph 4-6. Calibration procedures and the equipment necessary are presented in paragraph 4-14.

WARNING

Caution should be exercised when servicing this power supply. The metal shell of some of the electrolytic capacitors may be as much as 10, 100 volts above chassis ground. Capacitors C5 and C219 are high-quality, oil-filled units capable of retaining a charge for several days. Before servicing or removing tubes, all capacitors and plate caps should be shorted to the chassis. It is recommended that the shorting wire remain connected to C5 and C219 during servicing to prevent build-up of capacitor voltage due to dielectric absorption.

4-3. PREVENTIVE MAINTENANCE

- 4-4. Periodic cleaning of the Model 410B is desirable because of the high voltage present. Any contamination, particularly on the high voltage capacitors, may cause corona discharge, which will appear as noise in the output voltage. Components may be cleaned with Freon, or with a lint-free rag saturated with denatured alcohol. Ceramic switches may be cleaned with denatured alcohol or tetrachloroethane. After cleaning, the ceramic surface should be coated with a 10% solution of Dow Corning silicon fluid (200 viscosity grade). Denatured alcohol or tetrachloroethane may be used as a thinner for the silicon fluid.
- 4-5. Printed circuit boards in the 410B are coated with a polyurethane compound to prevent moisture absorbtion. If components are replaced, it will be necessary to recoat the circuit board in the immediate area of the part replaced. Aerosol containers of Epocast, available from Furane Plastics, Inc., Los Angeles, California, provide a convenient way of recoating the circuit board area. The recoated circuit board does not require baking.

4-6. TROUBLESHOOTING

4-7. It is recommended that all checks be made with the POLARITY switch turned to negative output polarity, and when possible, the output voltage set to 500 volts. Most voltages are referred to the positive bus, and when

the output polarity is negative, the positive bus is connected to the earth grounded chassis. This affords some protection to the person performing the tests.

4-8. The power supply should be allowed to warm-up before the following accuracy checks. Usually, one-half hour of operation is adequate. Figure 4-1 is a list of various failures and probable causes. Reference to Figure 4-1 will occasionally indicate the cause of a failure. Components may be located by referring to Section V. A list of equipment required is given in Figure 4-2.

4-9. AUXILIARY VOLTAGES

- a. Connect the 881A common lead to the 410B chassis ground (shell of the OUTPUT connector).
- b. Set the POLARITY switch to negative.
- c. Set the POWER switch to ON.
- d. Set the OUTPUT VOLTAGE controls to zero.
- e. Set the HIGH VOLTAGE switch to ON.
- f. Connect the 881A positive lead to the junction of CR206 and CR207.
- g. The 881A should indicate $+10 (\pm 1)$ vdc.
- h. Connect the 881A positive lead to the junction of CR205 and CR206.
- i. The 881A should indicate $+20 (\pm 2)$ vdc.
- j. Connect the 881A positive lead to the junction of CR205 and R203.
 - k. The 881A should indicate +95 (±9.5) vdc.
- 1. Connect the 881A positive lead to the junction of CR208 and R205.
 - m. The 881A should indicate -110 (\pm 11) vdc.
- n. Connect the 881A positive lead to the positive end of C211. Also connect the 910A RMS Voltmeter across the same points as the 881A.
- o. The 881A should indicate 125 (± 2.25) vdc. The 910A should indicate less than 0.005 volts rms ripple.

4-10. LINE REGULATION AND RIPPLE

a. Connect the variable transformer between the 410B and the line. Set the transformer to 115 volts output.

WARNING

A good power line ground must be provided when the variable transformer is used.

- b. Connect the 910A and the 881A across the OUTPUT connector. Also connect the oscilloscope across the OUTPUT connector.
 - c. Set the POWER switch to ON.
 - d. Set the HIGH VOLTAGE switch to ON.
- e. Set the OUTPUT VOLTAGE controls to 500 volts, and record the voltage indicated by the 881A.
- f. Increase the variable transformer output to 127 vac,

SYMPTOM	PROBABLE CAUSE	REMEDY
To output	Blown fuse	Check fuse F1, and replace if necessary.
, odepar	Open heater of V202	Determine if tube will warm-up; if not replace.
	Open R104; Shorted C219	Check and replace if necessary.
Constant percentage error	Out of calibration	Recalibrate per paragraph 4-14.
in output voltage	Defective R106	Check and replace if necessary.
Percentage error over part of range or Output erratic over part of range	Defective wirewound resistor in the sampling string. (R107 thru R140)	Set the output voltage to a maximum and decrease one switch position at a time until the error disappears. The defective resistor will be found at the last switch position in which the error was noted.
o t t annotic aron ontire	Defective V202	Test and replace if necessary.
Output erratic over entire range	Defective output voltage switch, or dirty switch printed circuit board.	Replace switch, clean printed circuit board.
	Defective R104 or R106	Check and replace if necessary.
Output rises to over 10KV and follows line voltage variations	Defective V202, or Q208 thru Q214	Test and replace if necessary.
variations	Open sampling resistor or switch	Test R107 thru R140, and S3 thru S6, and replace if necessary.
Poor load regulation	Defective Q208 thru Q214	Check and replace if necessary.
Output voltage suddenly rises above preset value	Internal arcing in V202	Check and replace if necessary.
Noise in output	Dirty high-voltage switches, capacitors, or printed circuit boards.	Clean per paragraph 4-4.
Excessive drift	Defective V201, Q213, or Q214	Check by replacement.
Excessive ripple	Excessive ripple in +125V auxiliary supply	If auxiliary supply ripple exceeds approximately 5 mv, replace defective component.
	Defective C219, C5, Q213, or Q214	Check and replace if necessary.
Loss of control	Open R243	. Check and replace if necessary.

Figure 4-1. TROUBLESHOOTING

EQUIPMENT	SPECIFICATIONS REQUIRED
Variable transformer	3 ampere capacity, 100 volts to 130 volts output.
RMS Voltmeter. Fluke Model 910A, or equivalent	RMS measurement of non-sinusoidal waves.
DC Differential Voltmeter, Fluke Model 881A, or equivalent	Voltmeter/Voltage Divider must be capable of measuring 0 to 10,000 vdc with a minimum accuracy of 0.05% + 50
Voltage Divider, Fluke Model 80E-10, or equivalent	microvolts. Voltmeter must have a 1 millivolt null detector for use with the 80E.
Oscilloscope, Tektronix Model 541, with Type L plug-in unit, or equivalent	Minimum sensitivity of 5 mv/cm. 10 MHz bandwidth.
Blocking capacitor	0.05 microfarads, 10,000 vde rating.
Load resistors	50K ±5%, 5W 1K ±5%, 2W
Insulated screwdriver	Must have sold insulated shaft capable of withstanding 10,000 vdc.

Figure 4-2. EQUIPMENT REQUIRED FOR TROUBLESHOOTING AND CALIBRATION

while observing the 910A. Ripple indicated by the 910A and by the oscilloscope should be less than 1 millivolt rms, and 5 millivolts peak-to-peak, respectively. The voltage change indicated by the 881A should be less than 5 millivolts.

- g. Decrease the transformer output from 115 vac to 103 vac, while observing the 910A. Ripple indicated by the 910A and by the oscilloscope should be less than 1 millivolt rms and 5 millivolts peak-to-peak, respectively. The voltage change indicated by the 881A should be less than 5 millivolts from step e.
- h. If desired, repeat steps e. through g. with the POLARITY switch in the opposite polarity.
- Set the variable transformer to 115 vac output, and set the HIGH VOLTAGE switch to STANDBY-RESET.
 - Disconnect the 881A and the 910A.
- k. Connect the 80E-10 Voltage Divider to the output of the 410B, and connect the 881A to the output of the voltage divider.
- 1. Connect the 10KV blocking capacitor in series with the 910A input, and connect the series combination across the 410B OUTPUT connector.
 - m. Set the HIGH VOLTAGE switch to ON.
- n. Set the OUTPUT VOLTAGE controls to 10,000 volts, and record the voltage measured by the voltmeter/voltage divider combination.
- o. Increase the variable transformer output to 127 vac, while observing the 910A. Ripple indicated by the 910A and by the oscilloscope should not exceed 1 millivolt rms and 5 millivolts peak-to-peak, respectively. The voltmeter/voltage divider should indicate less than 0.100 volts change in the 410B output voltage from the value measured in step n.
- p. Decrease the transformer output from 115 vac to

103 vac, while observing the 910A. Ripple indicated by the 910A and by the oscilloscope should not exceed 1 millivolt rms and 5 millivolts peak-to-peak, respectively. The voltmeter/voltage divider should indicate less than 0.100 volts change in the 410B output voltage from the value measured in step n.

q. If desired, repeat steps n. through p. with the POLARITY switch in the opposite polarity.

4-11. LOAD REGULATION

- a. Connect the variable transformer between the 410B and the line. Set the transformer to 103 vac output.
- b. Connect the 881A across the OUTPUT connector.
- c. Set the POWER switch to ON.
- d. Set the HIGH VOLTAGE switch to ON.
- e. Set the OUTPUT VOLTAGE controls to 500 volts.
- f. Record the voltage indicated by the 881A.
- g. Carefully connect the 50K load resistor across the OUTPUT connector. The load current should be 0.010 ampere.
- h. The 881A should indicate less than 0.005 volts change from the voltage measured in step f.

4-12. STABILITY

- a. Connect the 80E-10 to the output of the 410B, and connect the 881A to the output of the voltage divider.
 - b. Set the POWER switch to ON.
 - c. Set the HIGH VOLTAGE switch to ON.
 - d. Set the OUTPUT VOLTAGE controls to 10,000 volts.
- e. After a minimum operating time of 30 minutes, record the voltage indicated by the voltmeter/voltage divider.

f. After an additional 60 minutes of operation, the voltage change indicated by the voltmeter/voltage divider should be less than 0.5 volt (0.0005 volt indicated by the 10V tap).

4-13. OUTPUT AT ZERO VOLTS

- a. Connect the voltmeter to the OUTPUT connector.
- b. Set the POWER switch to ON.
- c. Set the OUTPUT VOLTAGE controls to zero.
- d. Set the HIGH VOLTAGE switch to ON.
- e. The 881A should indicate less than 0.030 volts in both output polarities. If not, exchange Q213 and/or Q214 with another transistor of the same type used elsewhere in the instrument, so that the beta of the two transistors is more nearly equal.

4-14. CALIBRATION

4-15. The Model 410B may be calibrated as often as necessary. However, it is recommended that the reference voltage be checked every 30 days, and the overcurrent limit be checked every six months. As reference tube V201 (83A1) ages, the reference voltage changes slightly. It has been found that some reference tubes are more stable than others. However, most reference tubes will change by less than 0.1% for every 100 hours of operation after the first hundred hours. If the output voltage is recorded before the reference current is adjusted, a more realistic calibration period can be determined. The equipment required for calibration is given in Figure 4-2.

4-16. REFERENCE CURRENT ADJUSTMENT

- a. Connect the 80E-10 Voltage Divider to the output of the 410B, and connect the voltmeter to the output of the voltage divider.
- b. Set the OUTPUT VOLTAGE controls to 9,900 volts, using only the first two dials.
- c. Set the POLARITY switch to negative.
- d. Set the POWER switch to ON.
- e. Set the HIGH VOLTAGE switch to ON.
- f. Set the differential voltmeter to differentially measure 9,900 volts, considering the voltage divider ratio. For example, using the 80E-10 Voltage Divider, set the differential voltmeter to measure 9.9 volts.

WARNING

Use an insulated screwdriver to adjust R104. The case of this resistor is 85 volts above

chassis ground for negative output polarity, and may be as much as 10,085 volts above ground for positive output polarity.

g. After a minimum operating time of 30 minutes, adjust R104 for a null on the differential voltmeter. Within 24 volts of 9,900 volts is sufficient (24 millivolts with the 80E-10). This control may be adjusted through the access hole on the right side panel near the front.

4-17. OVERCURRENT TRIP ADJUSTMENT

- a. Connect a 1K $\pm 5\%,~2W$ resistor to the OUTPUT connector.
 - b. Set the OUTPUT VOLTAGE controls to zero.
 - c. Set the POLARITY switch to negative.
 - d. Set the POWER switch to ON.
 - e. Set the HIGH VOLTAGE switch to ON.
- f. Set the 410B for maximum current trip by turning R151 completely clockwise. This control may be adjusted through the access hole on the left side of the instrument.
- g. Set the OUTPUT VOLTAGE controls to that output voltage which will deliver 12 milliamperes to the load resistor. For example, 12 volts will deliver approximately 12 milliamperes to a 1K resistor.

NOTE

If it is desired to set the overcurrent trip to some value other than 12 milliamperes, then the output voltage in step g. should be chosen to provide the desired trip current.

h. Slowly turn R151 counter-clockwise until the overcurrent trip removes power from the high voltage transformer.

4-18. METER CALIBRATION

- 4-19. The panel meter is adjusted at the factory, and will not usually require attention. If adjustment is necessary, proceed as follows:
 - a. Set the POWER switch to ON.
 - b. Set the HIGH VOLTAGE switch to ON.
- c. Set the OUTPUT VOLTAGE controls to 10,000 volts.
- d. Adjust R150 so that meter deflection corresponds to 10KV. This resistor may be adjusted through the access hole on the bottom of the instrument.

SECTION V

LIST OF REPLACEABLE PARTS

5-1. INTRODUCTION

- 5-2. This section contains information necessary to describe all normally replaceable parts. Separate assembly lists are used to describe the parts on the final assembly and various assemblies and subassemblies. Each list has a corresponding illustration on which the parts for that list are identified. Parts are called out on both lists and illustrations by reference designations from the schematic diagram. Those parts (mechanical) which have no reference designation are shown on the illustrations by Fluke stock number.
- 5-3. Each list provides the following information on each part:
- a. The REFERENCE DESIGNATION column indicates the reference designation used on the schematic diagram.
- b. The DESCRIPTION column describes the part in words, along with any applicable values, tolerances, etc. Indentation is used to show assembly, subassembly, and parts relationship. See abbreviations and symbols on next page.
- c. Entries in the FLUKE STOCK NUMBER column indicate the number by which Fluke stocks the part. This number should be used when ordering parts from the Fluke factory or your Fluke representative.
- d. Entries in the MFR. column indicate a typical manufacture of the part by the manufacturer's code number. Appendix A lists the manufacturers and their code numbers.
- e. Entries in the MFR. PART NO. column are part numbers assigned by the manufacturer indicated in the Mfg. column.
- f. The number in the TOTAL QTY column indicates the total quantity of the part used in the instrument. "REF" indicates that the total quantity of the part has been previously given. The total quantity of each part is listed the first time the part appears. All other listings of the same part refer back to the reference designation of the first appearance of the part for the total quantity.
- g. The number in the REC. QTY. column indicates the recommended spares quantity necessary to support

- approximately one to five instruments for a period of two years. The basis used to select the recommended spares quantity is that a small group of parts will be required to correct a majority of the problems that occur. Since there is a chance that any part may fail, a stock of at least one of every part used in addition to the recommended parts will be needed for complete maintenance during one year of isolated service.
- h. The USE CODE column identifies certain parts which have been added, deleted, or modified during production of the instrument. Each part for which a use code has been assigned may be identified with a particular instrument serial number by consulting the Use Code Effectivity List at the end of this section. These changes are normally made when improved components become available or when the latest circuit improvements are developed by our engineering department. The serial number listed indicates the instruments in which that particular part was used. The symbol "~" is used to indicate an approximate serial number. If a different part should be used for replacement, it is listed by Fluke stock number in the description column.

5-4. HOW TO OBTAIN PARTS

- 5-5. Standard components have been used whenever possible. Thus, most parts can be obtained locally. However, parts may be ordered directly from the manufacturer using the manufacturer's part number or from Fluke using the Fluke stock number. In addition, the most commonly replaced parts that can not be obtained locally may be obtained from your Fluke representative. If a part you have ordered has been replaced by a new or improved part, Fluke will normally send you this part along with an explanation.
- 5-6. When ordering parts from Fluke always include:
- a. Reference designation, description, and Fluke stock number.
- b. Instrument model and serial number.
- c. Most structural parts are not listed. In this case, give complete description, function, and location of part.

5-7. ABBREVIATIONS

5-8. PREFIX SYMBOLS

ac Al assy cap car flm cer comp conn cps db dc dpdt dpst elect fxd Ge gmv Hz K	alternating current aluminum assembly capacitor carbon film ceramic composition connector cycles per second decibel direct current double pole double throw double pole single throw electrolytic fixed germanium guaranteed minimum value hertz (cycles per second) kilohm	T G M K or k h da d c m u n	tera giga mega kilo hecto deka deci centi milli micro nano pico femto anto	10 ¹² 10 ⁹ 10 ⁶ 10 ³ 10 ² 10 10 ⁻¹ 10 ⁻² 10 ⁻³ 10 ⁻⁶ 10 ⁻⁹ 10 ⁻¹² 10 ⁻¹⁵ 10 ⁻¹⁸
kc or Kc	141-auglo			
kHz or KHz	kilocycle kilohertz (kilocycles per/sec) kilovolt			
kv kva	kilovoli-ampere			
ma	milliampere			
Mc or MC	megacycle megahertz (megacycles per/sec)			
MHz	megahertz (megaby	5-9. QUANT	TITY SYM BO	ols .
meg or M met flm	metal film	5-9,		
míg	manufacturer			ampere
m.v	millivolt	a or am	qı	farad
mw	milliwatt	f		henry
na	nanoampere picofarad	h hr		hour
pf	peak inverse voltage	Ω		ohm
piv plstc	plastic	sec		second volt
pp	neak-to-peak	v or V		watt
ppm	parts per million	w or W	Ī	W Prese
rect	rectifier resistor			
res	resistor root-mean-square			
rms	slow-blow			
sb Si	silicon			
s/N	serial number			
sw	switch			
spdt	single pole double throw single pole single throw			
spst	tontalum			
Ta tc	temperature coefficient			
tstr	transistor		TO NOTE	S AND SYMBOLS
ua	microampere	5-10. SPE	CIAL NOTE.	3 2
uf	microfarad			
uv	microvolt volt ampere	~ 0	ıΛΛ	Approximate serial
va	alternating current volts	\sim 0		number
vac var	variable			- I amban indicated
vdc	direct current volts	Use (0000-000000	Part number indicated should be used if re-
w	watt			placement is required.
wydc	direct current working volts			pracement to a train
ww	wirewound			

REF DESIG.	DESCRIPTION	FLUKE STOCK NO.	MFR.	MFR. PART NO.	TOT. QTY.		USE CODE
	Final Assembly (see Figure 5-1)	410B	89536				
	Front Panel Assembly (see Figure 5-2)	3158-175729 (410B-406)	89536	3158-175729	1		Andrea de la constanta de la c
	Switch PCB Assembly (see Figure 5-3)	1702-175679 (410B-401)	89536	1702-175679	1		***************************************
Account of the Contract of the	Amplifier PCB Assembly (see Figure 5-4)	1702-175687 (410B-402)	89536	1702-175687	1		***************************************
преченения сталаласть.	Time Delay PCB Assembly (see Figure 5-5)	1702-175695 (410B-403)	89536	1702-175695	1		
	Rectifier PCB Assembly (see Figure 5-6)	1702-175703 (410B-404)	89536	1702-175703	1		
C1, C2	Cap, cer, 0.01 uf $-20/+80\%$, $500V$ (not illustrated)	1501-105668	56289	29C9B5	2		
C3, C4	Cap, oil, $.25/.25$ uf $\pm 20\%$, 7.5 KV	1505 - 163766	99120	LK-75-504C	1		,
C5	Cap, oil, 0.25 uf $\pm 20\%$, 10KV	1505-163758	99120	LK-100-254N	1		
C219	Cap, oil, 0.05 uf ±10%, 10KV	1505-161158	01884	SMLE503-10M	1		
J2	Connector, H. V.	2104-100172	02660	MS 3102A -18 - 16S	2		
K301	Relay, time delay 115V, 30 sec.	4502-105288	70563	115N030T	1	1	
P1	Line cord set, 3 wire	6005-102822	70903	PVC-PH-70	1		i
R1	Res, comp, $470\Omega \pm 10\%$, 1W	4704-109710	01121	GB4711	1		
S3A	Switch section, rot., 11 positions, 1 pole	5107-176149	71590	PS-284	1		
S3B	Switch rotor assembly	5108-175612	895 3 6	5108-175612	1		
57	Switch, rotary	5110-175711	8 95 36	5110-175711	1		
S8, S9, S10	Switch, interlock (S9 and S10 not illustrated)	5104-115196	01963	E13-00A	3		
T1	Transformer, aux.	5600-176032	895 36	5600-176032	1		
Г2	Transformer, H. V.	5600-176024	89536	5600-176024	1		
V202	Electron tube, type 4-65A	5701-192351	89536	5701-192351	1	1	
	Meter cover	3155-169813	97945	1704-462-Y	1		
in difference for other	Feet, rubber (not illustrated)	2819-103309	83478	9102-W	4		
-	Knob assembly, pol. sw.	2405-175620	895 3 6	2405-175620	1		
The state of the s	Switch detent, 10 position, short	5108-155945	76854	239914-Н	3		
The second secon	Switch detent, 10 position, long	5108-167833	Ì	246305-Н	1	and marketing.	

REF DESIG.	DESCRIPTION	FLUKE STOCK NO.	MFR.	MFR. PART NO.	TOT. QTY.	REC. QTY.	USE CODE
DEGICA	Mating connector, H. V. Knob, 1-1/2" Cable clamp Cap, H.V. connector	2405 - 170035 2804 - 100149	89536 02660	MS3106A-18-16P 2405-170035 AN3057 .0 3155-172122	5 1 2		




Figure 5-1. FINAL ASSEMBLY (Sheet 1 of 2)




Figure 5-1, FINAL ASSEMBLY (Sheet 2 of 2)

REF DESIG.	DESCRIPTION	FLUKE STOCK NO.	MFR.	MFR. PART NO.	TOT. QTY.	REC. QTY.	USE CODE
	Front Panel Assembly	3158-175729 (410B-406)	89536	3158-175729	REF		
DSI	Lamp, red lens	3903-100206	91802	1040A1	2		
DS2	Lamp, white lens	3903-100214	91802	1040A4	1		
DS3	Lamp, red lens	3903-100206	91802	1040A1	REF		
Fi	Fuse, 3A, sb (not illustrated)	5101-109280	03614	Type MDA	1	3	
J1	Connector, H. V.	2104-100172	02660	MS3102A-18-	REF		
M1	Meter, 50Ω , $500-0-500$ ua $\pm 2\%$	2901-166231	89536	2901-166231	1		
S1	Switch, toggle, DPST, 8A, 250V	5106-114835	04009	81024-GB	1	***************************************	
S2	Switch, toggle, DPDT, 15A, 115V	5106-157883	73559	2GL50-63	1		
	Fuseholder	2102-100107	03614	нкр	1		
	Handle, 6-7/16"	2404-101584	05704	805	2		
	Cap, H.V. connector	3155-172122	89536	3155-172122	REF		




Figure 5-2. FRONT PANEL ASSEMBLY

REF	DESCRIPTION	FLUKE	MED	MFR.	TOT	REC	USE
DESIG.		STOCK NO.	MFR.	PART NO.	QTY.		CODE
TO THE PARTY OF TH	Switch PCB Assembly	1702-175679 (410B-401)	89536	1702-175679	REF		
C102	Cap, elect, $10 \text{ uf } -10/+75\%$, 150 V	1502-106351	56289	30D106G150DF	4 1	1	
R104	Res, var, WW, $1K \pm 20\%$, $1-1/4W$	4702-113266	71450	Type 110	1		
R105	Res, WW, $1.02K \pm 0.1\%$, $1/2W$	4707-145128	89536	4707-145128	1		A
R106 R106 R107	Res, WW, 40K $\pm 0.02\%$, 1/2W Res, WW, 41.6K $\pm 0.05\%$, 1 1/2W	4707 - 146530 4707 - 199778	1	4707-146530 4707-199778	1	1	A B
thru R124	Res, WW, 250K $\pm 0.05\%$, 2W	4707-156448	89536	4707-156448	18	2	
R125 thru R128	Res, WW, 100K ±0.1%, 1W	4707-142349	89536	4707-142349	4		
R129	Res, WW, $50K \pm 0.1\%$, $1W$	4707-156455	89536	4707-156455	1		
R130 thru R133	Res, WW, $10K \pm 0.1\%$, $1W$	4707-131664	89536	4707-131664	1		_
R134	Res, WW, $5K \pm 0.1\%$, $1/2W$	4707-149708	89536	4707-149708	1	· .	
R135 thru R138	Res, met flm, 1K ±1%, 1/2W	4705 - 151324	75042	Type CEC-TO	6		
R139	Res, met flm, $499\Omega \pm 1\%$, $1/2W$	4705-151514	75042	Type CEC-TO	1		
R140	Res, var, WW, $6000 \pm 10\%$, 2W	4702-155523	71450	Туре 252	1	į	
R141	Res, comp, $4.70 \pm 5\%$, 1W	4704-109785	01121	GB47 G 5	3		
R142 thru R148	Res, met flm, 2.67M $\pm 1\%$, 2W	4705-169391	14298	CM-2	7		
R149	Res, met flm, $309\Omega \pm 1\%$, $1/2W$	4705-172130	75042	Type CEC-TO	1		
R150, R151	Res, var, WW, 3K ±20%, 1-1/4W	4702-149781	71450	Type 110	2		-
R152	Res, comp, $82\Omega \pm 5\%$, $1/2W$	4704-108746	01121	EB8205	1		
R153	Res, comp, 4.7 Ω ±5%, 1W	47 04-109 7 85	01121	GB47G5	REF		
S4, S5, S6	Switch section, rot, 1 pole, 10 positions	5108-1559 7 8	76854	239956-FC	3		***************************************
					THE PROPERTY OF THE PROPERTY O		

Rev. 1




Figure 5-3. SWITCH CIRCUIT BOARD ASSEMBLY

REF DESIG.	DESCRIPTION	FLUKE STOCK NO.	MFR.	MFR. PART NO.	TOT.	REC.	USE CODE
	Amplifier PCB Assembly	1702-175687 (410B-402)	89536	1702-175687	REF		
C201	Cap, fxd, cer, 0.0033 uf $\pm 20\%$, 1KV	1501-106674	56289	5GA-D33	1		
C202, C203	Cap, fxd, elec, 40 uf -10/+100%, 500V	1502-106765	56289	Type 66D	2	1	
C204	Cap, fxd, elec, 200 uf $-10/+75\%$, 12V	1502-150284		30D207G012DF		1	
C205, C206	Cap, fxd, cer, 0.1 uf -20/+80%, 500V	1501-105684	56289	41C92	5	_	
C207	Cap, fxd, cer, 0.0047 uf $\pm 10\%$, 500V Cap, fxd, cer, 0.005 uf $\pm 20\%$, 1KV	1501~106724 1501-105650	1	CF-472 C023B102H502M	1 1		E F
C208, C209	Cap, fxd, cer, $0.1 \text{ uf } -20/+80\%$, 500 V	1501-105684	56289	41C92	REF		
C210	Cap, fxd, cer, $0.05 \text{ uf } -20/+80\%$, 500 V	1501-105676	56289	33C58	1		
C211	Cap, fxd, elect, $40 \text{ uf } -10/+75\%$, 250 V	1502-161273	56289	34D406G250GL4	1	1	
C212	Cap, fxd, cer, 0.01 uf $-20/+80\%$, 500V	1501-105668	56289	29C9B5	1		
C215	Cap, fxd, cer, 180 pf $\pm 10\%$, 500V	1501-105890	71590	BB60181KS3N	1		
C216	Cap, fxd, cer, 0.0027 uf GMV, 500V	1501-108211	72982	ED-, 0027	1		
C217	Cap, fxd, cer, $0.1 \text{ uf } -20/+80\%$, 500 V	1501-105684	58289	41C92	REF		
C218	Cap, fxd, cer, 0.01 uf GMV, 1.6KV	1501-106930	71590	DD16-103	1		
CR201 thru CR204	Diode, silicon, 600 PIV, 1A	4802-112383	05277	1N4822	10	2	
CR205	Diode, Si, zener, 75V, 0.005A	4803-168096	81483	1N3041A	1	1	
CR206, CR207	Diode, Si, zener, 10V, type 1N961A	480 3-11332 4	07910	1N961A	2	1	
CR208	Diode, Si, zener, 110V, 0.005A	4803-168104	81483	1N3045A	1	1	
CR209	Diode, silicon, 600 PIV, 1A	4802-112383	05277	1N4822	REF		Е
CR210, CR211	Diode, silicon, 100 PIV, 1A	4802-116111	05277	1N4822	4	1	
CR213, CR214	Diode, silicon, 100 PIV, 1A	4802-116111	05277	1N4822	REF		A Andrews
CR215	Diode, Si, 200 PIV, 12A	4802-188854	04713	MR1112	1	1	
CR216, CR217	Diode, silicon, 600 PIV, 1A	4802-112383	05277	1N4822	REF	***************************************	F
DS201 thru DS204	Lamp, neon, Type NE2E	3902-100347	71744	NE 2E	4	2	
Q201, Q202	Transistor, NPN, Si, Type 2N3053	4805-150359	95303	2N3053	2	1	

REF DESIG	DESCRIPTION	STOCK NO	MFR				USE CODE
Q203,	Transistor, NPN, Si	4805-203489	07910	CDQ10656	3	2	
Q204 Q205,	Transistor, NPN. Si	4805-203489	07910	CDQ10656	REF		С
Q206 Q205,	Transistor, NPN, Si	4819-168716	07263	S19254	2		D
Q206 Q208, Q209, Q210	Transistor, PNP, Si	4805-190389	04713	SM4144	3	1	C
Q208, Q209,	Transistor, PNP, Si	4805-159491	04713	SS7504	3		D
Q210 Q211,	Transistor, NPN, Si	4805-203489	07910		REF		
Q212 Q213, Q214	Transistor, NPN, Si	4805-203489			REF		C
Q214 Q213, Q214	Transistor, NPN, Si	4805-235812					D
R201	Res, comp, $10\Omega \pm 10\%$, $1/2W$	4704-108092	1		1		
R202, R203	Res, WW, 30K ±5%, 10W	4706-155433			2		
R204	Res, comp, $75\Omega \pm 5\%$, $1/2W$	4704-10875		1			
R205	Res, comp, 150K ±10%, 1W	4704-10980	1				С
R206 R206	Res, comp, 2.7K ±10%, 2W Res, comp, 2.2K ±10%, 2W	4704-11014 4704-10996	7 0112	HB2221	1		D
R207, R208	Res, comp, $6.2K \pm 5\%$, $1/2W$	4704-10862			7O 1		
R209	Res, met flm, $20K \pm 1\%$, $1/2W$ (not illustrated)	4705-16243					
R210	Res, met flm, 100K ±1%, 1/2W (not illustrated)	4705-1513				2	
R211	Res, comp, 47K ±10%, 1/2W	4704-1084	1			1	
R212	Res, comp, 2.7K ±5%, 1/2W	4704-1088 4705-1610	1		то	1	
R213	Res, met flm, 40.2K ±1%, 1/2W (not illustrated)					1	
R214	Res, met flm, 80.6K ±1%, 1/2W	4705-1506 4704-1084		Í		1	
R215	Res, comp, 180K ±10%, 1/2W	4704-1084	1			2	
R216	Res, comp, 82K ±10%, 1/2W	4704-1084	I	i		1	
R217	Res, comp, 39K ±10%, 1/2W (not illustrated)		į				
R218		4704-108	- 1		072	1	
R219 R219	and the second name	4707-172	∪/2 89 >	330 4/0/-1/2	ا سر با ل	•	

REF DESIG	DESCRIPTION	STOCK NO	MFR	MFR PART NO	i i	1	USE
Daga							
R220	Res, WW, 5K ±0.1%, 1/2W	4707-149708	89536	4707-149708	REF		A
R221 R221	Res, WW, $4K \pm 0.03\%$, $1/2W$ Factory selected part	4707-131672	89536	4707-131672	1	1	A B
R222	Res, comp, 10K ±10%, 1W	4704-109389	01121	GB1031	1		
R223, R224	Res, comp, $47\Omega \pm 10\%$, $1/2W$	4704-108688	01121	EB4701	2		
R226	Res, comp, $68K \pm 10\%$, $2W$	4704-110114	01121	HB6831			
R227	Res, comp, 1K ±10%, 1/2W	4704-108563	01121	EB1021	1		
R228, R229	Res, comp, 82K ±10%, 1/2W	4704-108498	01121	EB8231	REF		
R230	Res, comp, $270\Omega \pm 5\%$, $1/2W$	4704-159616	01121	EB2715	1		
R231	Res, comp, $680\Omega \pm 10\%$, $1/2W$	4704-108712	01121	EB6811	1		
R232	Res, comp, $47K \pm 10\%$, $1/2W$	4704-108480	01121	EB4731	REF		
R233	Res, comp, $4.7K \pm 10\%$, $1/2W$	4704-108381	01121	EB4721	2		
R234	Res, comp, $100K \pm 10\%$, $1/2W$	4704-108126	01121	EB1041	1		
R235	Res, comp, $4.7K \pm 10\%$, $1/2W$	4704-108381	01121	EB4721	REF		
R236	Res, met flm, $1K \pm 1\%$, $1/2W$	4705-151324	75042	Type CEC-TO	REF		
R237, R238	Res, comp, 270K $\pm 10\%$, 1/2W	4704-108258	01121	EB2741	2		
R239	Res, comp, $1.2M \pm 10\%$, $1/2W$	4704-108407	01121	EB1251	1		
R240	Res, comp, $2.7\Omega \pm 10\%$, 1W (not illustrated)	4704-109850	01121	GB27G1	1		
R241	Res, comp, $4.7\Omega \pm 5\%$, 1W	4704-109785 C	01121	EB47G5	REF		
R242	Res, comp, 2.2K ±10%, 1/2W		.	EB2221	1		
R243	Res, met flm, 1K±1%, 1/2W (not illustrated)		ľ	Type CEC-TO	1		
V201 V201	Electron Tube, volt. ref, Type 83A1 Factory selected	5701-170076	89536	5701-170076	*****	Virent	A B
Andreas de la constante de la							Managas de l'actività de l

1>>

 $R221,\,R219,\,and\,\,V201$ constitute a factory selected and matched set. For replacement, order part number 239707.




Figure 5-4. AMPLIFIER CIRCUIT BOARD ASSEMBLY

REF DESIG.	DESCRIPTION	FLUKE STOCK NO.	MFR.	MFR. PART NO.	TOT.	 USE CODE
	Time Delay PCB Assembly	1702-175695 (410B-403)	89536	1702-175695	REF	
CR301 thru CR304	Diode, silicon, 600 PIV, 1A	4802-112383	05277	IN4822	REF	
K302	Relay, armature, DPDT, 230 VAC, 5A	4504-148940	73949	A410-060713-00	1	
K303	Reed switch	4501-169698	89536	4501-169698	1	
R301	Res, WW, 1K ±5%, 10W	4706-157933	06136	10F-1000	1	




Figure 5-5. TIME DELAY CIRCUIT BOARD ASSEMBLY

5-11. USE CODE EFFECTIVITY

The following list of use codes is intended to allow the customer to determine the effectivity of all replaceable parts. All parts with no code are used on all instruments with serial numbers above 123. New codes will be added as necessary on forthcoming instruments.

mately 1331.	USE CODE	EFFECTIVITY
B Model 410B serial number 813 and on. C Model 410B serial number 123 thru 1072. D Model 410B serial number 1073 and on. E Model 410B serial number 123 thru approx mately 1331. F Model 410B serial number approximatel		Model 410B serial number 123 and on.
C Model 410B serial number 123 thru 1072. D Model 410B serial number 1073 and on. E Model 410B serial number 123 thru approx mately 1331. F Model 410B serial number approximatel	Α	Model 410B serial number 123 thru 812.
D Model 410B serial number 1073 and on. E Model 410B serial number 123 thru approx mately 1331. F Model 410B serial number approximatel	В	Model 410B serial number 813 and on.
 E Model 410B serial number 123 thru approx mately 1331. F Model 410B serial number approximatel 	С	Model 410B serial number 123 thru 1072.
mately 1331. F Model 410B serial number approximatel	D	Model 410B serial number 1073 and on.
	E	Model 410B serial number 123 thru approximately 1331.
	F	Model 410B serial number approximately 1332 and on.

Section 7 General Information

7-1. This section of the manual contains generalized user information as well as supplemental information to the List of Replaceable Parts contained in Section 5.

REV.6 11/87

List of Abbreviations and Symbols

A or amp	ampere	hf	high frequency	(+) or pos	positive
ac	alternating current	Hz	hertz	pot	potentiometer
af	audio frequency	IC	integrated circuit	р-р	peak-to-peak
a/di	analog-to-digital	if	intermediate frequency	ppm	parts per million
assy	assembly	in	inch(es)	PROM	programmabile read-only
AWG	american wire gauge	intl	internal		memory
В	bel	1/0	mput/output	psi	pound-force per square inc
bcd	binary coded decimal	k	kilo (10 ³)	RAM	random-access memory
°C	Celsius	kHz	kilohertz	र्त	radio frequency
cap	capacitor	kΩ	kilohm(s)	rms	root mean square
ccw	counterclockwise	kV	kilovolt(s)	ROM	read-only memory
cer	ceramic	11	low frequency	s or sec	second (time)
cermet	ceramic to metal(seal)	LED	light-emitting diode	scope	oscilloscope
ckt	circuit	LSB	least significant bit	SH	shield
cm	centimeter	LSD	least significant digit	SI	silicon
cmrr	common mode rejection ratio	M	mega (10 ⁶)	serno	serial number
comp	composition	m	mitti (10 *3)	sr	shift register
cont	continue	mA	milliampere(s)	Ta	tantalum
crt	cathode-ray tube	max	maximum	lb	terminal board
cw	clockwise	mf	metal film	tc	temperature coefficient or
d/a	digital-to-analog	MHz	megahertz		temperature compensating
dac	digital-to-analog converter	min	minimum	texo	temperature compensated
dB	decibel	mm	millimeter		crystal oscillator
de	direct current	ms	millisecond	tp	test point
dmm	digital multimeter	MSB	most significant bit	u or μ	micro (10 ⁻⁶)
dvm	digital voltmeter	MSD	most significant digit	uhf	ultra high frequency
elect	electrolytic	MTBF	mean time between failures	us or μ s	microsecond(s) (10 ⁻⁶)
ext	external	MTTR	mean time to repair	uut	unit under test
F	farad	mV	millivolt(s)	٧	volt
og:	Fahrenheit	mv	multivibrator	٧	voltage
FET	Field-effect transistor	МΩ	megohm(s)	var	variable
ff	flip-flop	n	nano (10-*)	VCO	voltage controlled oscillato
freq	frequency	na	not applicable	vhf	very high frequency
FSN	federal stock number	NC	normally closed	vif	very tow frequency
g	gram	(-) or neg	negative	w	watt(s)
G	giga (10°)	NO	normally open	ww	wire wound
gd	guard	ns	nanosecond	xlmr	transformer
Ge	germanium	opni ampi	operational amplifier	xstr	transistor
GHz	gigahertz	p	pico (10*12)	xtal	crystal
gmv	guaranteed minimum value	para	paragraph	xtio	crystal oscillator
gnd	ground	pcb	printed circuit board	Ω	ohm(s)
H	henry	pF	picofarad	μ	micro (10 ⁻⁶)
hdi	heavy duty	pn pn	part number	•	

D9816

Westermann Wilhelm Augusta-Anlage Mannheim-Nackarau Germany

Marcon Electronics Corp Kearny, New Jersey

Nytronics Comp. Group Inc. Darrlingon, South Carolina

Welwyn International Inc. Westlake, Ohio

Acrovox Corp.

New Bedford, Massachusetts

Film Capacitors Inc. Passaic, New Jersey

AMP, Inc.

Harrisburg, Pennsylvania

Allen Bradley Co. Milwaukec, Wisconsin

TRW Electronics & Defense Sector Lawndale, California

Texas Instruments Inc. Semiconductor Group Dallas,Texas

01537

Motorola Communications & Electronics Inc.

Franklin Park, Illinois

RCL Electronics/Shallcross Inc. Electro Compononts Div. Manchester, New Hampshire

01884

Sprague Electric Co. (Now 56289)

Varian Associates Inc. Pulse Engineering Div. Convoy, Connecticut

Spectrol Electronics Corp. City of Industry, California

Amperex Electronic Corp. Ferrox Cube Div. Saugenties, New York

General Instrument Corp. Government AVX Corp.

Systems Div. Westwood, Massachusetts

Sonar Radio Corp. Hollywood, Florida 02533

Leigh Instruments Ltd. Frequency Control Div. Don Mills, Ontario, Canada

02606 Fenwal Labs

Division of Travonal Labs Morton Grove, Illinois

0266

Bunker Ramo-Eltra Corp. Amphonol NA Div. Broadview, Illinois

02735

RCA-Solid State Div. Somerville, New Jersey

Arco Electronics Inc. Chatsworth, California

03508

Goneral Electric Co.

Semiconductor Products & Batteries Aubum, New York

03797

Genisco Technology Corp.

Eltronics Div. Rancho Dominquez, Calif.

Gilbert Engineering Co.Inc Incon Sub of Transitron Electronic Corp. Glondale, Arizona

03888

KDI Electronics Inc. Pyrofilm Div. Whippany, New Jersey

03911 Clairex Corp.

Clairex Electronics Div. Mount Vemon, New York

03980

Muirhead Inc. Mountainside, New Jersey

Cooper Industries, Inc. Arrow Hart Div. Hartford, Coonecticut

Essex International Inc. Wire & Cable Div. Anaheim, California

04221

Midland-Ross Corp. Midtex Div. N. Mankato, Minnesota

04222

AVX Ceramics Div. Myrtle Beach, S. Carolina

04423

Telonic Berkley Inc. Laguna Beach, California 04713

Motorola Inc. Serriconductor Group Phoenix, Arizona

05236

Jonathan Mfg. Co. Fullerton, California

05245

Corcom Inc. Libertyville, Illinois

05276 ITT Pomona Electronics Div. Pomona, California

05277

Westinghouse Elec. Corp. Semiconductor Div. Youngwood, Pennsylvania

05397

Union Carbide Corp. Materials Systems Div. Cleveland, Ohio

05571

Sprague Electric Co. (Now 56289)

05574

Viking Connectors Inc. Sub of Criton Corp. Chatsworth, Calif.

EG & G Wakefield Engineering Wakefield, Massachusetts

05972

Locute Corp. Newington, Connecticut

General Electric Co. Electric Capacitor Product Section Columbia, S. Carolina

Fairchild Weston Systems Inc. Data Systems Div. Sarasota, Florida

06192

La Deau Mfg. Co. Glendale, California

06229

Electrovert Inc. Elmsford, New York

06383

Panduit Corp. Tinley Park Illinois

06473

Bunker Ramo Corp. Amphonol NA Div. SAMS Operation Chatsworth, California

Beede Electrical Instrument Penacook, New Hampshire

06665

Precision Monolithics Sub of Bourns Inc. Santa Clara, California

06666

General Devices Co. Inc. Indianapolis, Indiana

06739

Electron Corp. Littleton, Colorado

06743 Gould Inc. Foil Div. Eastlake, Ohio

06751

Components Inc. Semoor Div. Phoenix, Arizona

06776

Robinson Nugent Inc. New Albany, Indiana

06915

Richco Plastic Co. Chicago, Illinois

06961

Vernitron Corp. Piezo Electric Div. Bedford, Ohio

06980

Varian Associates Inc. Eimac Div. San Carlos, California

07047

Ross Milton Co., The Southampton, Penna.

Westinghouse Electric Corp. Industrial & Government Tube Div. Horseheads, New York

07233

Benchmark Technology Inc. City of Industry, Calif.

07239

Biddle Instruments Blue Bell, Penna.

07256

Silicon Transistor Corp. Sub of BBF Inc. Chelmsford, Massachusetts

07261 Avnet Corp.

Culver City, California

Fairchild Camera & Instrument Semiconductor Div. Mountain View, California

07344

Bircher Co. Inc., The Rochester, New York

07557 Campion Co. Inc. Philadelphia, Penna.

07597 Burndy Corp. Tape/Cable Div. Rochester, New York

07716 TRW Inc. (Can use 11502) IRC Fixed Resistors/ Burlington Burlington, lowa

07792 Lerma Engineering Corp. Northampton, Massachusetts

07810 Bock Corp. Madison, Wisconsin

07933 Raytheon Co. Semiconductor Div. Mountain View, Calif.

08235 Industro Transistor Corp. Long Island City, New York

08261 Spectra-Strip An Eltra Co. Garden Grove, Calif.

08530 Reliance Mica Corp. Brooklyn, New York

08718 ITT Cannon Electric Phoenix Div. Phoenix, Arizona

08806 General Electric Co. Minature Lamp Products Cleveland, Ohio

08863 Nylomatic Fallsington, Penna.

08988 Skottie Electronics Inc. Archbald, Pennsylvania

09021 Airco Inc. Airco Electronics Bradford, Penna.

Cornell-Dublier Electronics Fuquay-Varina, N. Carolina

09214
General Electric Co.
Semiconductor Products Dept.
Auburn, New York

09353 C and K Components Inc. Newton, Massachusetts 09423 Scientific Components Inc. Santa Barbara, California

O9579 CTS of Canada, Ltd Streetsville, Ontario

09922 Burndy Corp. Norwalk, Connecticut

09969 Dale Electronics Inc. Yankton, South Dakota

09975 Burroughs Corp. Electronics Components Detroit, Michigan

10059 Barker Engineering Corp. Kenilworth, New Jersey

10389
Illinois Tool Works Inc.
Licon Div.
Chicago, Illinois

10582 CTS of Asheville Skyland, N. Carolina

11236 CTS Corp. Berne Div. Berne, Indiana

11237 CTS Corp of California Paso Robles Div. Paso Robles, California

11295 ECM Motor Co. Schaumburg, Illinois

11358
Columbia Broadcasting System
CBS Electronic Div.
Newburyport, Massachusetts

Vacuum Can Co.Best Coffee Maker Div. Chicago, Illinois

11502
TRW Inc.
TRW Resistive Products Div.
Roome North Carolina

Boone, North Carolina
11503

Keystone Columbia Inc. Freemont, Indiana

Telodyne Relays Teledyne Industries Inc. Hawthome, California 11711 General Instrument Corp. Rectifier Div. Hicksville, New York

11726 Qualidyne Corp. Santa Clara, Califomia

12014 Chicago Rivet & Machine Co. Naperville, Illinois

12040 National Semiconductor Corp. Danbury, Connecticut

12060 Diodes Inc. Northridge, California

12136
PHC Industries Inc.
Formerly Philadelphia Handle Co.
Camden, New Jersey

12300 AMF Canada Ltd. Potter Brumfield Guelph, Ontario, Canada

12323 Practical Automation Inc. Shelton, Connecticut

12327 Freeway Corp. Cleveland, Ohio

12443
Budd Co., The
Plastics Products Div.
Phoenixville, Pennsylvania

12581 Hitachi Metals Inemational Ltd. Hitachi Magna Lock Div. Big Rapids, Missouri

12615 US Terminals Inc. Cincinnati, Ohio

12617 Hamlin Inc. Lake Mills, Wisconsin

Clarostat Mfg. Co. Inc. Dover, New Hampshire

12/49 James Electronic Inc. Chicago, Illinois

MicroMetals Inc. Anaheim, California

12881 Metex Corp. Edison, New Jersey

12895 Cleveland Electric Motor Co. Cleveland, Ohio 12954 Microsemi Corp. Components Group Scottsdale, Arizona

12969 Unitrode Corp. Lexington, Massachusetts

13050 Potter Co. Wesson, Mississippi

13103 Thermalloy Co., Inc. Dallas, Texas

13321 Solitron Devices Inc. Tappan, New York

13511 Bunker-Ramo Corp. Amphenol Cadre Div. Los Gatos, California

13606 Sprague Electric Co. (Use 56289)

13689 SPS Technologies Inc. Hatfield, Pennsylvania

13919 Burr-Brown Research Corp. Tueson, Arizona

14099 Semtech Corp. Newbury Park, California

14140 McGray-Edison Co. Commercial Development Div. Manchester, New Hampshire

14193 Cal-R-Inc. Santa Monica, California

14298 American Components Inc. an Insileo Co. RPC Div. Conshohocken, Pennsylvania

14298 ACIC Inc. Sub of Insileo Corp. Research Triangle Park, NC

Wells Electronics Inc. South Bend, Indiana

14482 Watkins-Johnson Co. Palo Alto, California

14552 Microsemi Corp. Santa Ana, California

14655
Comell-Dublier Electronics
Div. of Federal Pacific
Electric Co. Govt Cont Dept.
Newark, New Jersey

14704 Crydom Controls (Division of Int Rectifier) It Segundo, California

14752 Electro Cube Inc. San Gabriel, California

14936 General Instrument Corp. Discrete Semi Conductor Div. Hicksville, New York

14949 Trompeter Electronics Chatsworth, California

15412 Amtron Midlothian, Illinois

15542 Scientific Components Corp. Mini-Circuits Laboratory Div. Brooklyn, New York

15636 Elec-Trol Inc. Saugus, California

15782 Bausch & Lomb Inc. Graphics & Control Div. Austin, Texas

15801 Fenwal Eletronics Inc. Div. of Kidde Inc. Framingham, Massachusetts

15818
Teledyne Inc. Co.
Teledyne Semiconductor Div.
Mountain View, California

15849 Useco Inc. (Now 88245)

15898 International Business Machines Corp. Essex Junction, Vermont

16245 Conap Inc. Olean, New York

16258 Space-Lok Inc. Burbank, California

16352 Codi Corp. Linden, New Jersey

16469 MCL Inc. LaGrange, Illinois

The Hart Control of the Control of t

16473
Cambridge Scientific Industries
Div. of Chemed Corp.
Cambridge, Maryland

16733 Cablewave Systems Inc. North Haven, Connecticut

16742 Paramount Plastics Fabricators Inc. Downey, California

16758 General Motors Corp. Delco Electronics Div. Kokomo, Indiana

17069 Circuit Structures Lab Burbank, California

17117 Electronic Molding Corp. Woonsocket, Rhode Island

17338 High Pressure Eng. Co. Inc. Oklahoma City, Oklahoma

17545 Atlantic Semiconductors Inc. Asbury Park, New Jersey

17745 Angstrohm Precision, Inc. Hagerstown, Maryland

17856 Siliconix Inc. Santa Clara, California

18178 E G & Gyactee Inc. St. Louis, Missouri

18324 Signetics Corp. Sacramento, California

18520 Sharp Electronics Corp. Paramus, New Jersey

18542 Wabash Inc. Wabash Relay & Electronics Div. Wabash, Indiana

18565 Chomerics Inc. Woburn, Massachusetts

18612 Vishay Intertechnology Inc. Vishay Resistor Products Group Malvem, Pennsylvania

18632 Norton-Chemplast Santa Monica, California

18677 Scanbe Mfg. Co. Div. of Zero Corp. El Monte, California

18736 Voltronics Corp. East Hanover, New Jersey 18927 GTE Products Corp. Precision Material Products Business Parts Div. Timsville, Pennsylvania

19315 Bendix Corp., The Navigation & Control Group Terboro, New Jersey

19451 Perine Machinery & Supply Co., Kent, Washington

19613
Minnesota Mining & Mfg. Co.
Textool Products Dept.
Electronic Product Div.
Irving, Texas

19647 Caddock Electronics Inc. Riverside, California

19701 Mepco/Centralab Inc. A N. American Philips Co. Mineral Wells, Texas

20584 Enochs Mfg, Inc. Indianapolis, Indiana

20891 Cosar Corp. Dallas, Texas

21317 Electronics Applications Co. El Monte, California

21604 Buckeye Stamping Co. Columbus, Ohio

21845 Solitron Devices Inc. Semiconductor Group Rivera Beach, Florida

22526
DuPont, El DeNemours & Co. Inc.
DuPont Connector Systems
Advanced Products Div.
New Cumberland, Pennsylvania

22767 ITT Semiconductors Palo Alto, California

22784 Palmer Inc. Cleveland, Ohio

23050 Product Comp. Corp. Mount Vernon, New York

23732 Tracor Applied Sciences Inc. Rockville, Maryland

23880 Stanford Applied Engineering Santa Clara, California 23936 William J. Purdy Co. Pamotor Div. Burlingame, California

24347
Penn Engineering Co.
S. El Monte, California

24355 Analog Devices Inc. Norwood, Massachusetts

24444 General Semiconductor Industries, Inc. Tempe, Arizona

24655 Genrad Inc. Concord, Massachusetts

24759Lenox Fugle Electronics Inc. South Plainfield, New Jersey

24796 AMF Inc. Potter & Brumfield Div. San Juan Capistrano, Calif.

24931 Specialty Connector Co. Greenwood, Indiana

25088 Siemen Corp. Isilen, New Jersey

25099 Cascade Gasket Kent, Washington

25403
Amperex Electronic Corp.
Semiconductor & Micro-Circuit Div.
Slatersville, Rhode Island

25706 Dabum Electronic & Cable Corp. Norwood, New Jersey

26629 Frequency Sources Inc. Sources Div. Chelmsford, Massachusetts

26806 American Zettler Inc. Irvine, California

27014 National Semiconductor Corp. Santa Clara, California

27167 Coming Glass Works Coming Electronics Wilmington, North Carolina

27264 Molex Inc. Lisle, Illinois

27440 Industrial Screw Products Los Angeles, California

27745 Associated Spring Barnes Group Inc. Syracuse, New York

Relcom (Now 14482)

Positronic Industries Springfield, Missouri

Minnesota Mining & Mfg. Co. Consumer Products Div.

3M Center

Saint Paul, Minnesota

Serv-O-Link Euless, Texas

28478

Deltrol Corporation Deltrol Controls Div. Milwaukee, Wisconsin

Hewlett Packard Co. Corporate HQ Palo Alto, California

Emerson Electric Co. Gearmaster Div. McHenry, Illinois

Heyco Molded Products Kenilworth, New Jersey

Monsanto Co. Santa Clara, California

Stackpole Components Co. Raleigh, North Carolina

Omega Engineering Inc. Stamford, Connnecticut

Jolo Industries Inc. Garden Grove, California

Symbex Corp. Painesville, Ohio

AB Enterprise Inc. Ahoskie, North Carolina

Asvid Engineering Inc. Laconia, New Hampshire

Itron Corp.

San Diego, California

Illinois Tool Works Inc. Chicago, Illinois

General Instrument Corp. Capacitor Div. Hicksville, New York

Solid State Scientific Inc. Willow Grove, Pennsylvania

Alpha Industries Inc. Microelectronics Div. Hatfield, Pennsylvania

Metro Supply Company Sacramento, California

Army Safeguard Logistics Command Huntsville, Alabama

31746

Cannon Electric Woodbury, Tennessee

Budwig

Ramona, California

31918 IIT-Schadow Eden Prairio, Minnesota

32293 Intervil

Cupertino, California

Mura Corp.

Westbury, Long Island, N.Y.

32559

Santa Ana, California

Griffith Plastics Corp. Burlingame, California

Advanced Mechanical Components Northridge, California

Murata Erie North America Inc. Carlisle Operations

Carlisle, Pennsylvania

Bourns Inc. Trimpot Div. Riverside, California

Colorado Crystal Corp. Loveland, Co'orado

General Electric Co. Owensboro, Kentucky

Epoxy Technology Inc. Billerica, Massachusetta

NEC Electronics USA Inc. Electronic Arrays Inc. Div. Mountain View, California

Nortek Inc.

Cranston, Rhode Island

Silicon General Inc. Garden Grove, California

34225

Advanced Micro Devices Sunnyvale, California

34359

Minnesota Mining & Mfg. Co. Commercial Office Supply Div. Saint Paul, Minnesota

34371

Harris Corp. Harris Semiconductor Products Group Melbourne, Florida

Intel Corp.

Santa Clara, California

34802

Electromotive Inc. Kenilworth, New Jersey

Hartwell Special Products Placentia, California

Renfrew Electric Co. Ltd. IRC Div. Toronto, Ontazio, Canada

36665

Mitel Corp. Kanata, Ontario, Canada

Mallory Capacitor Corp. Sub of Emhant Industries Indianapolis, Indiana

39003

Maxim Industries Middleboro, Massachusetts

Roderstein Electronics Inc. Statesville, North Carolina

42498

National Radio Melrose, Massachusetts

Nytronics Inc.(Now 53342)

Ohmite Mfg. Co. Skokie, Illinois

RCA Corp. New York, New York

Thomson Components-Mostek Corp. Carrollton, Texas

50120

49956

Raytheon Company

Lexington, Massachusetts

Executive Offices

Eagle-Picher Industries Inc. Electronics Div.
Colorado Springs, Colorado

50157

Midwest Components Inc. Muskegon, Mississippi

Hypertronics Corp. Hudson, Massachusetts

Litronix Inc. Cupertino, California

Aries Electronics Inc. Frenchtown, New Jersey

51372

Verbatim Corp. Sunnyvale, California

Murata Erie, No. America Inc. (Also see 72982) Marietta, Georgia

51499

Amtron Corp. Boston, Massachusents

CODI Semiconductor Inc. Kenilworth, New Jersey

Centre Engineering Inc. State College, Pennsylvania

Statek Corp. Orange, California

NEC America Inc. Falls Church, Virginia

Exar Integrated Systems Sunnyvale, California

52072

Circuit Assembly Corp. Irvine, California

Minnesota Mining & Mfg. Saint Paul, Minnesota

API Electronics

Haugpauge Long Island, New York

54590 58104 64155 Communication Systems RCA Com. Simeo Linear Technology Piscataway, New Jersey Electronic Components Div. Atlanta, Georgia Milpitas, California Cherry Hill, New Jersey 58474 Space-Lok Inc. Superior Electric Co. West M G Co. Lerco Div. American Gage & Machine Co. Bristol, Connecticut San Francisco, Calif. Burbank, California Simpson Electric Co. Div. Elgin, Illinois 59124 52531 KOA-Speer Electronics Inc. Sangamo Weston Inc. Hitachi Magnetics Bradford, Pennsylvania Weston Instruments Div. Plessey Capacitors Inc. (Now 60935) Edmore, Missouri Newark, New Jersey 59640 52745 55261 Supertex Inc. Rohm Corp & Whatney Timco LSI Computer Systems Inc. Melville, New York Sunnyvale, California Irvine, California Los Angeles, California 55285 Tusonix Inc. Evox Inc. Stettner-Electronics Inc. Beroquist Co. Tucson, Arizona Bannockburn, Illinois Chattanooga, Tennessee Minneapolis, Minnesota 66150 Thomas and Beus Corp. Entron Inc. Sprague-Goodman Electronics Inc. Syneriek Winslow Teltronics Div, Iowa City, Iowa Garden City Park, New York Santa Clara, California Glendale, New York Semtronics Corp. 52771 Michicon/America/Corp. Watchung, New Jersey Bering Industries Moniterm Corp. Schaumburg, Illinois Fremont, California Amatrom Div. 60395 Santa Clara, California 56282 Xicor Inc. Utek Systems Inc. Milpitas, California Almetal Universal Joint Co. Olathe, Kansas Cleveland, Ohio Western Digital Corp. 60399 Costa Mesa, California Torin Engineered Blowers Div. of Clevepak Corp. 56289 Atlantic India Rubber Works Inc. Sprague Electric Co. Torrington, Connecticut Chicago, Illinois Sangamo Weston Inc. North Adams, Massachusents (S∞ 06141) 56365 60705 Amporite Company 53217 Square D Co. Cera-Mite Corp. Union City, New Jersey Technical Wire Products Inc. Corporate Offices Palatine, Illinois (formerly Sprague) Grafton, Wisconsin Santa Barbara, California Belden Corp. Geneva, Illinois Opt Industries Inc. DAL Industries Inc. Westlake Capacitor Inc. Phillipsburg, New Jersey Wescorp Div. Mountain View, California Tantalum Div. Greencastle, Indiana 71002 53944 Bimbach Co. Inc. Glow-Lite 61804 Pauls Valley, Oklahoma Farmingdale, New York Shugart Associates M/A Com Inc. Sub of Xerox Corp. Sunnyvale, California Burlington, Massachusetts 71034 Bliley Electric Co. Shallcross Inc. Erie, Pennsylvania SAN-O Industrial Corp. Smithfield, North Carolina Zilog Inc. Bohemia, Long Island, NY Campbell, California Westinghouse Electric Corp. Sullins Electronic Corp. Bryant Div. Schurter Inc. Bridgeport, Connecticut San Marcos, California Vamistor Corp. of Tennessee Sevierville, Tennessee Petaloma, California Matsushita Electric Corp. Bussman Manufacturing Apple Rubber (Panasonic) Div. McGraw-Edison Co. Magnetics Inc. Lancaster, New York St. Louis, Missouri Secaucus, New Jersey Baltimore, Maryland Lear Siegler Inc. CTS Corp. TDK Endicott Coil Co. Inc. Energy Products Div. Garden City, New York Elkhart, Indiana Binghamton, New York Santa Ana, California Piher International Corp. ITT Cannon Div. of ITT Gates Energy Products Ward Leonard Electric Co.Inc. Arlington Heights, Illinois Fountain Valley, California Denver, Ohio Mount Vernon, New York 54937 64154 General Instrument Corp. DeYoung Mfg. Hitachi Magnalock Corp. Lamb Industries Clare Div. Chicago, Illinois Bellevue, Washington

Portland, Oregon

(Now 12581)

BEFORE CALCULATION

71590
Mepco/Centralab
A North American Philips Co.

Fort Dodge, Iowa

71707 Coto Corp. Providence, Rhode Island

General Instrument Corp. Lamp Div/Worldwide Chicago, Illinois

71785 TRW Inc, Cinch Connector Div, Eik Grove Village, Illinois

71984 Dow Coming Corp. Midland, Michigan

72005 AMAX Specialty Metals Corp.

Newark, New Jersey 72136

Electro Motive Mfg. Corp. Florence, South Carolina

AMCA International Corp. Continental Screw Div. New Bedford, Massachusetts

72259 Nytronies Inc. New York, New York

72619
Amperex Electronic Corp.
Dialight Div.
Brooklyn, New York

72653 G C Electronics Co. Div. of Hydrometals Inc. Rockford, Illinois

72794 Dzus Fastner Co. Inc. West Islip, New York

72928 Gulton Industries Inc. Gudeman Div. Chicago, Illinois

72982 Murata Erie N. America Inc. Erie, Pennsylvania

73138

Beckman Industrial corp.
Helipot Div.
Fullerton, California

73168 Fenwal Inc. Ashland, Massachusetts

73293 Hughes Aircraft Co. Electron Dynamics Div. Torrance, California 73445 Amperex Electronic Corp. Hicksville, New York

73559 Carlings witch Inc. Hanford, Connecticut

73586 Circle F Industries Trenton, New Jersey

73734
Federal Screw Products Inc.
Chicago, Illinois

73743
Fischer Special Mfg. Co.
Cold Spring, Kentucky

73893 Microdot Mt. Clemens, Mississippi

73899
JFD Electronic Components
Div. of Murata Erie
Oceanside, New York

73905 FL Industries Inc. San Jose, California

73949 Guardian Electric Mfg. Co. Chicago, Illinois

74199 Quam Nichols Co. Chicago, Illinois

74217 Radio Switch Co. Marlboro, New Jersey

74306 Piezo Crystal Co. Div. of PPA Industries Inc. Carlisie, Pennsylvania

74542 Hoyt Elect.Instr. Works Inc. Penacook, New Hampshire

74840 Illinois Capacitor Inc. Lincolnwood, Illinois

74970 Johnson EF Co. Waseca, Minnesota

75042 TRW In c. IRC Fixed Resistors Philadelphia, Pennsylvania

75297 Litton Systems Kester Solder Div. Chicago, Illinois

75376 Kurz-Kasch Inc. Dayton, Ohio 75378 CTS Knights Inc. Sandwich, Illinois

Kulka Electric Corp. (Now 83330) Mount Vernon, New York

75915 Tracor Littlefuse Des Plaines, Illinois

76854 Oak Switch Systems Inc. Crystal Lake, Illinois

77122
TRW Assemblies & Fasteners Group
Fastener Div.
Moutainside, New Jersey

77342 AMF Inc. Potter & Brumfield Div. Princeton, Indiana

77542 Ray-O-Vac Corp Madison, Wisconsin

77638
General Instrument Corp.
Rectifier Div.
Brooklyn, New York

77900 Shakeproof Lock Washer Co. (Now 78189)

Rubbereraft Corp. of CA Ltd. Torrance, California

78189 Illinois Tool Works Inc. Shakeproof Div. Elgin, Illinois

78277 Sigma Instruments Inc. South Braintree, Mass.

78290 Struthers Dunn Inc. Pitman, New Jersey

78553 Eaton Corp. Engineered Fastener Div. Cleveland, Ohio

78592 Stoeger Industries South Hackensack, New Jersey

79136 Waldes Kohinoor Inc. Long Island City, New York

79497 Western Rubber Co. Goshen, Indiana 79727 C - W Industries Southampton, Pennsylvania

79963 Zierick Mfg. Corp. Mount Kisco, New York

Tektronix Beaverton, Oregon

80009

80031 Mepco/Electra Inc. Morristown, New Jersey

Sousze & Communications Corp.
Western Development
Laboratories Div.
Palo Alto, California

80145 LFE Corp. Process Control Div. Clinton, Ohio

80183 Sprague Products (Now 56289)

80294
Bourns Instruments Inc.
Riverside, California

80583 Hammerlund Mfg. Co. Inc. Paramus, New Jersey

80640 Computer Products Inc. Stevens-Arnold Div. South Boston, Mass.

81073 Grayhill Inc. La Grange, Illinois

81312 Litton Systems Inc. Winchester Electronics Div. Watertown, Connecticut

81439 Therm-O-Disc Inc. Mansfield, Ohio

International Rectifier Corp.
Los Angeles, California

81590 Korry Electronics Inc. Scattle, Washington

Chicago Lock Co. Chicago, Illinois

82227 Airpax Corp. Cheshire Div. Cheshire, Connecticut

82240 Simmons Fastner Corp. Albany, New York

82305

Palmer Flectronics Corp. South Gate, California

87700

Switchcraft Inc. Sub of Raytheon Co. Chicago, Illinois

82415 Airpax Corp Frederick Div. Frederick, Maryland

82872 Roanwell Corp, New York, New York

82877 Rotron Inc. Custom Div. Woodstock, New York

82879 ITT Royal Electric Div. Pawtucket, Rhode Island

83003 Varo Inc. Garland, Texas

83014 Hanwell Corp. Placentia, California

83055 Signalite Fuse Co. (Now 71744)

83058 TRW Assemblies & Fasteners Group Fasteners Div. Cambridge, Massachusetts

83259 Parker-Hannifin Corp. O-Seal Div. Culver City, California

83298 Bendix Corp. Electric & Fluid Power Div. Eatonville, New Jersey

83315 Hubbell Corp. Mundelein, Illinois

Kulka Smith Inc.
A North American Philips Co.
Manasquan, New Jersey

Rubbercraft Corp. of America West Haven, Connecticut

83553 Asseciated Spring Barnes Group Gardena, California

83740 Union Carbide Corp. Battery Products Div. Danbury, Connecticut

THE PROPERTY TO SPECIAL PROPERTY.

84171 Arco Electronics Commack, New York

84411 American Shizuki TRW Capacitors Div. Ogallala, Nebraska

84613 FIC Corp. Rockville, Maryland

84682 Essex Group Inc, Peabody, Massachusetts

85367 Bearing Distributing Co. San Fransisco, California

85372 Bearing Sales Co. Los Angeles, California

85480 W. H. Brady Co. Industrial Product Milwaukee, Wisconsin

85932 Electro Film Inc. Valencia, California

86577
Precision Metal Products Co.
Peabody, Massachusetts

86684 Radio Corp. of America (Now 54590)

86928 Seastrom Mfg. Co. Inc. Glendale, California

87034 Illuminated Products Inc. (Now 76854)

88219 GNB Inc. Industrial Battery Div. Langhome, Pennsylvania

88245 Winchester Electronics Litton Systems-Useco Div. Van Nuys, California

88486 Triangle PWC Inc. Jewitt City, Connecticut

88690 Essex Group Inc. Wire Assembly Div. Dearborn, Michigan

89020 Amerace Corp. Buchanan Crimptool Products Div. Union, New Jersey

89265 Potter-Brumfield (Sec 77342) 89536 John Fluke Mfg. Co., Inc. Everett, Washington

89597 Fredericks Co. Huntingdon Valley, Penna.

89709
Bunker Ramo-Eltra Corp.
Amphenol Div.
Broadview, Illinois

89730 General Electric Lamp Div, Newark, New Jersey

90201
Mallory Capacitor Co.
Sub of Emhart Industries Inc.
Indianapolis, Indiana

90215 Best Stamp & Mfg. Co. Kansas City, Missouri

90303 Duracell Inc. Technical Sales & Marketing Bethel, Connecticut

91094
Essex Group Inc.
Suflex/IWP Div.
Nowmarket, New Hampshire

91247 Illinois Transformer Co. Chicago, Illinois

91293 Johanson Mig. Co. Boonton, New Jersey

91462 Alpha Industries Inc. Logansport, Indiana

Associated Machine Santa Clara, California

91506 Augat Inc. Attleboro, Massachusetts

91507 Froeliger Machine Tool Co. Stockton, California

91637 Dale Electronics Inc. Columbus, Nebraska

91662 Elco Corp. A Gulf Western Mfg. Co. Connector Div. Huntingdon, Pennsylvania

91737 ITT Cannon/Gremar (Now 08718) 91802 Industrial Devices Inc. Edgewater, New Jersey

91833 Keystone Electronics Corp. New York, New York

King's Electronics Co. Inc. Tuckahoe, New York

91929 Honeywell Inc. Micro Switch Div. Freeport, Illinois

91934 Miller Electric Co. Woonsocket, Rhode Island

91984 Maida Development Co. Hampton, Virginia

Norwalk Valve Co. S. Norwalk, Connecticut

92914 Alpha Wire Corp. Elizabeth, New Jersey

93332 Sylvania Electric Products Semiconductor Products Div. Wobum, Massachusetts

94144 Raytheon Co. Microwave & Power Tube Div. Quincy, Massachuseus

94222 Southco Iric. Concordville, Pennsylvania

94988 Wagner Electric Corp. Sub of Mcgraw-Edison Co. Whippany, New Jersey

95146 Alco Electronic Products Inc. Switch Div. North Andover, Massachusetts

95263 Leecraft Mfg. Co. Long Island City, New York 95275

Vitramon Inc. Bridgeport, Connecticut 95303

95303 RCA Corp. Receiving Tube Div. Cincinnati, Ohio

95348 Gordo's Corp. Bloomfield, New Jersey

95354 Methode Mfg. Corp. Rolling Meadows, Illinois

95573

Campion Laboratories Inc. Detroit, Michigan

95712 Bendix Corp. Electrical Comp. Div. Franklin, Indiana

95987 Weckesser Co. Inc. (Now 85480)

96733 SFE Technologies San Fernando, California

96853 Gulton Industries Inc. Measurement & Controls Div. Manchester, New Hampshire

96881 Thomson Industries Inc. Port Washington, New York

97525 EECO Inc. Santa Ana, California 97540

Whitehall Electronics Corp. Master Mobile Mounts Div. Fort Meyers, Florida

97913 Industrial Electronic Hardware Corp. New York, New York

97945 Pennwalt Corp, SS White Industrial Products Piscataway, New Jersey

97966 CBS Electronic Div. Danvers, Massachusetts

98094 Machlett Laboratories Inc. Santa Barbara, California

98159 Rubber-Teck Inc. Gardena, California 98278

Malco A Microdot Co. South Pasadena, California

98291 Sealectro Corp. BICC Electronics Trumbill, Connecticut

98372

Royal Industries Inc.(Now 62793)

98388 Lear Siegler Inc. Accurate Products Div. San Deigo, Califomia

99120 Plastic Capacitors Inc. Chicago, Illinois

99217 Bell Industries Inc. Elect, Distributor Div. Sunnyvale, California 99378 ATLIE of Delaware Inc.

N. Andover, Massachusetts


Mepco/Electra Inc. Roxboro Div. Roxboro, North Carolina

99515 Electron Products Inc. Div. of American Capacitors Duarte, California

99779 Bunker Ramo- Eltra Corp. Barnes Div. Lansdown, Pennsylvania

99800 American Precision Industries Delevan Div. East Aurora, New York

99942 Mepco/Centralab A North American Phillips Co. Milwaukee, Wisconsin



. N

U.S. SALES OFFICE AREAS

AL, Huntsville

John Fluke Mfa. Co. inc. 4920J Corporate Drive Huntsville, AL 35805-6202 (205) 837-0581

AZ, Phoenix

John Fluke Mfg. Co., inc. 2211 S. 48th Street Suite B Tempe, AZ 85282 (602) 438-8314

Tucson

(602) 790-9881

CA Southern

John Fluke Mfg. Co., Inc. P.O Box 19676 Irvine, CA 92713-9676 16969 Von Karman Stute 100 Irvine, CA 92714 (714) 863-9031

Burbank

John Fluke Mfg. Co., Inc. 2020 N. Lincoln Street Burbank, CA 91504 (213) 849-7181

Northern

John Fluke Mfg. Co., Inc. 2300 Walsh Ave., Bldg. K Santa Clara, CA 95051 (408) 727-0513

San Diego

(61.9) 292-7656

CO, Denver

John Fluke Mfg. Co., Inc. 14180 East Evans Ave. Aurora, CO 80014 (303) 695-1000

CT, Hartford

John Fluke Mfg. Co., Inc. Glen Locken East 41-C New London Turnpike Glastonbury, CT 06033 (203) 659-3541

DC, Washington (301) 770-1570

FL, Clearwater (813) 799-0087

Miami

(305) 462-1380

Orlando

John Fluke Mfg, Co.,Inc. 940 N. Fern Creek Ave. Orlando, FL 32803 (305) 896-4881

Tampa

(813) 251-9211

John Fluke Mfg. Co. Inc. 2700 Delk Rd , Suite 150 Marietta, GA 30067 (404) 953-4747

IL, Chicago

John Fluke Mfg. Co. Inc. 1150 West Euclid Ave Palatine, IL 60067 (312) 705-0500

IN. Indianapolis

John Fluke Mfg. Co. Inc. 8777 Purdue Rd. Suite 101 Indianapolis, IN 46268 (317) 875-7870

LA, Baton Rouge

(504) 924-1203

New Orleans

(504) 455-0814

MA, Boston

John Fluke Mfg. Co., Inc. Middlesex Tech Center 900 Middlesex Turnpike Building 8 Billerica, MA 01821 (617) 663-2400

MD, Baltimore

(301) 792-7060

Rockville

John Fluke Mfg. Co., Inc. 5640 Fishers Lane Rockville, MD 20852 (301) 770-1570

MI, Detroit

John Fluke Mfg. Co., Inc. 33031 Schoolcraft Livonia, MI 48150 (313) 522-9140

MN, Bloomington

John Fluke Mfg. Co., Inc. 1801 E. 79th St., Suite 9 Bloomington, MN 55420 (612) 854-5526

MO, St. Louis

John Fluke Mfg. Co., Inc. 11756 Borman Drive Suite 160 St. Louis, MO 63146 (314) 993-3805

NC. Greensboro

John Fluke Mfg. Co., Inc 1310 Beaman Place Greensboro, NC 27408 (919) 273-1918

NJ. Paramus

John Fluke Mfg. Co., Inc. PO Box 930 Paramus, NJ 07653-0930 West 75 Century Road Paramus, NJ 07652 (201) 262-9550

NM, Albuquerque

(505) 881-3550

NY, Rochester

John Fluke Mfg. Co., Inc. 4515 Culver Road Rochester, NY 14622 (716) 323-1400

OH, Cleveland

John Fluke Mfg. Co., Inc. 7830 Freeway Circle Middleburg Heights, OH 44130 (216) 234-4540

Dayton

John Fluke Mfg. Co., Inc. 5450 Far Hills Avenue Suite 111 Kettering, OH 45429 (513) 436-2224

OK, Northeast

(918) 749-0190

OR, Portland

(503) 227-2042

PA, Philadelphia

John Fluke Mfg. Co., Inc. 200 Lindenwood Drive Malvern, PA 19355 (215) 647-9550

Pittsburgh

(412) 261-5171

TX. Austin

(512) 459-3344

Dallas

John Fluke Mfg. Co., Inc. 1801 Royal Lane Suite 307 Dallas, TX 75229 (214) 869-0311

El Paso

(915) 533-3508

Houston

(713) 240-5995

San Antonio

John Fluke Mfg. Co., Inc. 10417 Gulfdale San Antonio, TX 78216 (512) 340-0498

UT, Salt Lake City 1801) 268-9331

WA. Seattle

John Fluke Mfg Co. Inc. 5020 148th Ave N.E. Suite #110 Redmond, WA 98052 (206) 881-6966


U.S. Government Specialists

(301) 770-1570, MD (203) 837-0581, AL (714) 863-9031, CA (301) 770-1570, MD USAF (513) 436-2224, OH (512) 340-2621, TX Security (301) 770-1570, MD

Service Center Areas

CA, Burbank (213) 849-4641 CA, Santa Clara (408) 727-0513 CO, Denver (303) 695-1000 FL, Orlando (305) 896-4881 IL, Chicago (312) 705-0500 MD, Rockville (301) 770-1576 NJ, Paramus (201) 262-9550 TX, Dallas (214) 869-2848 WA, Everett (206) 356-5560

For more information on Fluke products or Sales Offices you may dial (800) 426-0361 toll-free in most of the U.S.A. From Alaska, Hawaii, or Washington phone (206) 356-5400. From Canada and other countries phone (206) 356-5500.



INTERNATIONAL SALES OFFICES

Crices S.A. Virrey del Pino 4071 DPTO E-65 1430 CAP FED Bunnos Astes, Argentino Fet 54-1-552-5248 TLX (390) 22284 COASN AR

Fruke Asia Ltd Shiin Tak Centre, Room 15itt 200 Connaught Road Certiat Hong Kong Tel 852 5 482116 TLX (780) 87058 FLUKE FAX (852) 5-479863

Anetratia 4

Elmeasco Instruments Pty. Ltd. P O Box 30 Concord, NSW 2137 Australia Tel: 61-2-736-2888 TLX (790) 25887 A/B ELSCOAA 25887 FAX 61-2-733663

Elmeasco Instruments Pty, Ltd P.O. Box 623 12 Maroondah Highway Ringwood, Victoria 3134 Australia Tel: 61-3-879-2322 TLX: (790) 30418 A/B: ELTENTAA 30418 FAX. (61) (3) 879-4310

Elmeasco Instruments Pty, Ltd. P.O. Box 274 Salisbury, Queenstand 4107 Australia Tel: 61-7-875-1444

TLX: (790) 44062 A/B: ELMQLDAA44062

Elmeasco Instruments Pty, Ltd. P O. Box 154 Prospect, South Australia 5082 Tel: 61-8-344-9000 TLX: (790) A/6-87519 Etmeesco Instruments Pty, Ltd.

P.Q. Box 413 Scott House 46-48 Kings Park Road West Perth, Western Australia 6005 Australia Tel: 61-9-481-1500 TLX: (790) 94765 A/B: ASECSAA 94765

Austria # Welter Rekirsch Elektronische Gerate GmbH & Co. Vertrieb KG Obachgasse 28 1220 Vienne, Austrie Tel; 43-222-25-36-26 TLX; (847) 134759 FAX: 43-222-25-72-75

FAX (61) (9) 322-2075

Bahraín 🕈 Basma W.L.L P.O. Box 5701 Manama, Bahrain Tel: 973-251364; TLX: (955) 9003 FAX: (965) 245218

Bangladesh • Motherland Corporation 24 Hatkhola Road, Tikatuli Dacca-3, Bangladesh Tel: 257249 TLX: (950) 642022 PCO BJ Cable: "MOTHERLAND" DACCA

Belgium a Industrial & Electro-Acoustic Systems 5600 MD Eindhoven The Netherlands Tel: 040-785539 Tix. 35000 phtc ni Fax: 040-785651

Botivia ●

Coasin Bolivia S.R.L. Casilla 7295 La Paz, Bolivia Tel: 591-2-40962 TLX. (336) 3233 COALAP BV Cabte, COALAP

ATP of Test Detroment tele Al Amazonas 422 Alphavilte, 06400 Barueri Sie Paulo Brazil Tel: 55: 11-421-5477 TEX: (391) 1171413 HtTK 8R

Brune: •

Rank O'Connoi's SDN BHD So 8 Blk D Sufir Shophouse Complex Mite 1 Jalan Tutong Bandar Seri Begawan Negara Brunei Datussalam Tel 673-2-23\09 or 23557 TLX (799) BU 2265 RANKOC

Canada •

Flicke Electronics Canada toc. 101, 1144 - 29th Avenue N.E. Calgary, Alberta T2E 7P1 Canada Tel: (403) 291-5215 Fax (403) 291-5219

Flicke Electronics Canada Inc. 400 Britannia Road East Unit #1 Mississauga, Ontario L4Z 1X9 Tel: (416) 890-7600 Fax: (416) 890-6866

Fluke Electronics Canada Inc 1265 Trans Canada Hwy Suite 130 Dorval, Quebec H9P 2V4 Ceneda Tel: (514) 685-0022 TLX (514) 685-0039

Fluke Electronics Caneda Inc. 1690 Woodward Drive Suite 216 Ottawa, Ontario K2C 3R8 Canada Tel: (613) 723-9453 Fax: (613) 723-9458

Chife .

Intronica Chile, Ltda. Casilla 16228 (Mail) Santlego 9, Chile Tel: 56-2-2321886 TLX: (332) 346351 INTRON CK

China, Peoples Republic of • Fluke International Corporation P.O. Box 9085 Beijing

People's Republic of China Tel: 86-01-65-7281 TLX: (716) 222237 FBSC CN

Instrimpex - Fluke Service Center 57, Xisi Dong De Jie Xicheng qu Beiling Peoples Republic of China

Tel: 86-01-65-7281

Colombia • Sistemas E Instrumentacion, Ltda. Carrera 13, No. 37-43, Ol. 401 Ap. Aereo 29583 Bogota DE, Colombia Tel: 57-232-4532 TLX: (396) 45787 COASN CO

Cyprus = Chris Radiovision, Ltd. P.O. Box 1989

Nicosia, Cyprus Tel: 357-21-66121; TLX: (826) 2395 Cyprus, Northern #

Ucok Buroteknik 2C & 2D Muftu Ziyai Street Lefkosa, Northein Cyprus Mersin 10, Turkey Tel: 90-741-357-20-71777 TLX (821) 57267 Czechoslovakia a

Amtest Associates, Ltd. Amtest House 75-79 Guildford Street Chertsey, Suirey KT16 9AS England Tet 44-9328-68355 TLX (851) 928855 AMTEST G Fax 44-9328-61919

Denmark • Tarpe Oliver Avis Ballerup Byver 322 2750 Ballerup Denmark

Lat. 45, 2, 558111 TEX (855) 35293 TOAS DK

Eastern European Countries w Amtest Associates, Lld.

Anitest House 75-19 Guidford Street Cherisey Surrey KT16 9AS England

Tel: 44-9328-68355 TLX (851) 928855 AMTEST G Fax 44-9028-61919

German Branch Office Amtest Associates, Lld Zugspitzstrasse 2A P (1) Box 1107 8011 Vaterstetten

West Germany Ter 49-81-067117, TLX; (841) 528332

Ecuador •

Proteco Coasin Cia., Ltda. P.O. Box 228-A Ave. t2 de Octubre 2285 y Oretlana Quito, Equador Tel: 593-2-529684 TLX (393) 22085 ESIND

Egypt Electronic Engineering Liaison Office P.O. Box 2891 Horreya 11361 Heliopolis, Calro

Egypt Tel: 20-2-695705, TLX: (927) 22782 England

N.V. Philips Industriel & Electro-Acoustic Systems 5600 MD Eindhoven The Netherlands Tel: 040-785539 Tix: 35000 ohte ni Fax: 040-785651

FAX: 679-314379

Awa New Zeeland Limited 37 Freeston Roed Walu Bay, P.O. Box 858 Suva, Fiji Tel: 679-312079, TLX: (792) FJ2347

Finland # Instrumenterium Elektroniike P.O. Box 64 02631 Espoo 63

Finland Tel: 358-0-5281 TLX: (857) 124426 HAVUL SF FAX: (358) 0-5021073 Teletex: (857) 8-100155 INSTRUE

France • M.B Electronique S.A. 606, Rue Fourney P.O Box 31 78530 BUC, France Tel: 33-1-39-56-81-31 TLX. (842) 695414 Fax: (33) (1) 3956-53-44

German Branch Office Ainlest Associates, Ltd. Zugspitzstrasse 2A P.O. Box 1107 8011 Vaterstetten West Germany Tet 49-81-067117 TLX: (841) 528332

Germany, Weel N.V Philips Industrial & Etectro-Acoustic Systems 5600 MD Eindhoven

The Netherlands Tel 040-785539 Tix: 35000 phtc nl Fax: 040-785651

Greece Hellenic Scientific Representations Ltd. Vrassida Street Athens 612, Greece Tet 30-1-7211140, TLX: (863) 219330

Hong Kong • Sensudi & Co (H.K.), Ed. 14th Floor, Great Engle Centre 23 Harbour Road Warrhai, Hong Kong Tal. 852-5183301222 TLX (780) 74766 SCHMC HX FAX 852-5 8918754

India •

Hindition Services Pvt. Ltd. 697A L. Jagmohandas Marg Bombay 400 006, india Tel: 91-22-8121316, 8125344 TLX (953) 1175326 HSPL tN

Bangatore Office Hinditron Services Pvt., Ltd 8th Main Road 33/44A Raj Mahai Vilas Extension Dangalore 560 080, India Tel: 91-812-363139 TLX (953) 08452741 Cable, TEKHIND BANGALORE

Calcutta Office Hinditron Services Pvt. Ltd. 5th Ftoor, Castle House 5/1A, Hungerford St. Calcutta 700 017, India Tel. 91-33-432628 TLX: (953) 214153

New Delhi Sales Hinditron Services Pvt. Ltd 204-206 Hemkunt Tower 98 Nehru Plece New Delhi, 110019, India Tel: 91 (11) 6410380 or 6414493 TLX: (953) 3161458 HSPL IN

New Delhi Service Hinditron Services Pvt., Ltd. Field Service Center 15, Community Centre Panchshita Perk New Delhi: 110 017, India Tet: 91 011 6433675

Cable: Tekcestre Delhi Hyderabad Office Hinditron Services Pvt. Ltd. Fleid Servica Canter Emerald Complex, 5th Floor 1.7.264 114 Sarorini Devi Road Secunderabad 500 003, India Tel: 91 842 821117 TLX: (953) 04256973 HSPL IN

indonesia • P.T. Lemda Trigune P.O. Box 6/JATJG Jekarta, 13001 Indonesia Tet: 62 21 8195365 TLX: (796) 46171 LAYARIA

Arma Baynelmetell Co., Ltd.

P.O. Box 951570 Pasdaran-Tebran Tel: 98-21-248717, TLX: (951) 213648 ireland

N.V. Phillips Industrial & Electro-Acoustic Systems 5600 MD Eindhoven The Netherlands Tel. 040+785539 Tix: 35000 phtc ni Fax: 040-785651

tsraei a R D.T. Etectronics Engineering Ltd. P.O. Box 43137 Tei Avlv 61430 Israei Tei: (3) 972-3-483211 TLX (922) 371452 RDT (L Fax: 972-3-492190

Italy = Sistrel S.p.A. Via le Erminio Spalla 41 00142 Rome, Italy Tet: 39-6-504-1367, TLX: (843) 625857 Fax (39) 6-504137



Milan Office Sistref Sin A Via Pelizza da Volpedo 59 20092 Cinisello Balsamo Milan, Italy Tel: 39-2-6161893, TLX (843) 334643 Fax (39) 2-6182440

Naples Office Sistrel S p.A V-a Cintia Parco S Paolo 35 80126 Naples, Italy Tel 39-81-7678700 Fax (39) 81-7661361

Japan •

John Fjuke Mig Co. inn Japan Branch Sumitomo Higashi Shinbashi Bida 1-1-11 Hamamatsucho Minato-ku, Tokyo 105, Japan Tel: 81-3-434-0181 TLX (781) 2424331 FLUK JPJ FAX: 81-3-434-0170

Osaka Sales Office John Fluke Mfg. Co., Inc. Japan Branch Katsushige Building 2-45 Kohraibashi Higashi ku, Osaka 541 Japan Tel. 81-6-229-0871 FAX: 81-6-229-1098

Korea, Republic of • Myoung Corporation Yeo Eui Do P.O. Box 14 Seoul, Korea Tel: 82-2-784-9942 MYOUNG TLX: (787) K24283 FAX: (82) 2-784-2387

Kuwait =

Al Bahar International Group P.O. Box 26672 Safat 13127 Safat, Kuwait Kuwait, Arabian Gulf Tel: 965-848601, 847598 TLX: (959) 44822

Lebanon and Jordan # Mabek (Electronics Division) P.O. Box 13-5657 Beirul, Lebanon Tel: 812523 TLX (923) 22889 LIBANK LE

Mataysia ● Mecomb Malaysia Sdn Bhd P O. Box 24 46700 Petaling Jaya, Selangor, Malaysia Tel: 60-3-774-3422 TLX: (784) MA37764 MECOMB Fax: (6) 03-774-3414

Malta . Fabian Enterprises 20, Msida Road Gzira, Malta Tel: 513283/40216, TLX: (838) 1837

Mexico •

Mexicana de Electronica Industrial, S.A. (Mexel) Diagonal No. 27 Entre Calle de Eugenia Y Ave Colonia del Valle C.P. 03100, Mexico Tel: (905) 660-4323 TLX: (383) 1771038 FAIRME

Executone De Monterrey, S.A. Ave Gonzalitos NTE 545 Monterrey N.L., Mexico Tel: 90-5-480400, 472625 TLX: (383) 382659

Morocco e Oussama S.A. Angle Boulevard Emile Zola et Rue Rethel P.O. Box 2007 Casa Casablanca Morocco Tel: 212-24-13-38, TLX: (933) 28879 M

Nepal . Associated Enterprises GPO Box 790, Pyaphai Tole Kathmandu, Nepal Tel: 13868, TLX: (947) 2568 (ASOENT NP) Netherlands

N v Philips Industrial & Flectio Acoustic Systems 5600 MD Eindhoven The Netherlands Let 940 785539 Tix 35000 phic or Eax 040-185651

New Zealand •

Auckland Office
Northrop Instruments & Systems, Etd. 459 Khyber Pass Road Private Bag, Newmarket Arckland t New Zealand Tel 64-9-501-801, 501-219 TEX (191) 21570 FAX 64-9-543430

Wellington Office Northrop Instruments & Systems Ltd Information Technology Group First Floor, Northrop Bldg. 189-191 Willis Street P.O. Box 2406 Wellington, New Zealand Tel: 64-4-856-658 TLX. (791) 3380

Christohurch Office Northrop Instruments & Systems Ltd. Information Technology Group 110 Mandeville Street P.O. Box 8388 Christchurch, New Zealand Tel: 64-3-488-874 TLX: (791) 4801

Norway . Morgenstierne & Co A/S Konghallegate 3 P.O. Box 6688. Rodelokka Oslo 5, Norway Tel: (2) 356110, TLX: (856) 71719

Oman 🛮 OHI Telecommunications LLC P.O. Box 889 Muscat, Oman Tel: 968-603606 TLX: (926) 5052 TELECOM ON Pakistan #

International Operations (PAK), Ltd. 505 Muhammadi House Ll. Chundrigar Road P.O. Box 5323, Karachi, Pakistan Tel: 92-21-221127, TLX: (952) 24494 PIO PK

Importaciones y Representaciones Electronicas S.A. Avda, Franklin D. Roosevelt 105 Lima 1, Peru Tel: 51-14-28-8650 TLX: (394) 25663 PE IREING

Philippines, Republic of Spark Radio & Elactronics, Inc. Greenhills P.O. Box 610 San Juan, Metro Manila, Zip 3113 Philippines Tel: 63-2-775192, 704096 TLX: (722 or 732) 27901 BLA PH

Portugal =

Decada Espectral Equipamentos de Electronica Av. Bombeiros Voluntarios Lote 102B, Miraflores/Alges 1495 Lisbon, Portugal Tel: 351-1-4103420, TLX: (832) 15515

Romania
Amtest Associates Ltd. Amtest House 75-79 Guildford Street Chertsey, Surrey KT16 9AS Tel: 44-9328-68355, TLX: (851) 928855 FAX: 44-9328-61919

Saudi Arabia ... Electronic Equipment Marketing Co. P.O. Box 3750 Riyadh, Saudi Arabia Tel: 966-1-477-1650 TLX: (928) 401120 ZUHAIR SJ

Scotland N / Pheips Industrial & Lieutin Acoustic Systems 5600 MD Emillioner The Netherlands Tel 040 195539 Ex 35000 phtc nt Fax 040-785651 Singapore Republic of • Rank O'Connor's (PTE) Ltd

O Cornor House 98 Pasir Paniang Road Singapore 0511 Regublic of Simpapore Tel. 65-4237944-TLX-(786) RS21023 OCONSIN FAX 172450H CABLE CINECOM SINGAPORE

South Africa • Protea PNI PO Box 39127 Bramley 2018 Republic of South Africa Tel 27-11-786-3647 TLX. (960) 4-24409 SA FAX 27-11-786-1807

Spain 🛎 Equipos y Sistemas S.A. C/Apolonio Morales, 13-B Madrid 16, Spain Tel: 34-1-458-0150 TLX. (831) 42856 EYS E

Sri Lanka • Computerlink Data Systems, Ltd. 331 Union Place Colombo 2, Sri Lanka Tel: 94-1-28641 TLX: (954) 22455 COLINK CE

Sweden N.V Philips Industrial & Electro-Acoustic Systems 5600 MD Eindhoven The Netherlands Tel: 040-785539 Tix: 35000 phtc nl Fax: 040-785651

Switzerland a Traco Electronic AG Jenatschstrasse 1 8002 Zurich Switzerland Tel. 41-1-201-0711 TLX. (845) 815570 TROOCH

Syria : Mabek (Electronics Division) P.O. Box 4238 Damascus, Syria

Taiwan •

Schmidt Electronics Corp. Schmidt Electronics Corp. 5th FI, Cathay Min Sheng Commercial Bullding, 344 Min Sheng East Road Taipel Talwan R.O.C. Tel: 886-2-501-3468 TLX: (785) 11111 SCHMIDT Fax: (886) 2-502-9692

Thailand • Measuretronix Lld, 2102/63 Ramkamhaeng Rd. Huamark Bangkok 10240 Thailand Tel: 66 (2) 3742516, 3741632 TLX: (788) 82796 HUAMARK TH

Tunisia ■ Selep S A.R I... 6, Rue de Sparte Tunis - 1000 RP Tunisia Tel: 216-1-248093, TLX: (934) 13030

Turkey m Erkman Elektronik Aletter Ticaret Anonim Sirkeli Necatibey Cad 92/3 Karakoy, Islanbul, Turkey Tel 90 (11) 441546, TLX: (821) 24399 United Arab Emirates = Haris Al-Afag, Ltd PO Box 814f Kamai Hainza Bido Traffic Police Road Pubai U.A.E. Tel 911-4-283625 TLX (958) 48168 AFAQEM Al-Sanani Cen Trad Est.

PO Box 7187 Abur Dhabi U A E. Tel 971-2-821370, TLX (958) 23966

Uruguay • Coasin Uruguaya S.A. Casilla de Correo 1480 Libertaid 2529 Montevideo Uruguay Tel 598-2-789204 789015 TLX (398) UY23010 COAUR

USSR Amtest Associates Ltd. Amtest House 75-79 Guildlord Street Chertsey Surrey KT 16 9AS England Tel: 44-9328-68355, TLX. (851) 928855 FAX: 44-9328-61919

Venezuela • Coasin C.A. Calle 9 Con Calle 4, Edif Edinurbi Apartado de Correos NR-70,136 Los Ruices Caracas 1070-A, Venezuela Tel 58 (2) 241 · 03 · 09 TLX. (395) 21027 EMVEN VC

Yugosiavia = Amtest Associates Ltd. Amlest House 75-79 Guildford Street Chertsey, Surray KT16 9AS England Tel: 44-9328-68355, TLX: (851) 928855 FAX: 44-9328-81919

s Supplied and Supported by Fluka (Holland) B.V P.O. Box 2269 5600 CG Eindhoven Tha Netherlands Tel: (040) 45805, TLX: (844) 51846 FAX 31-40-457515

 Supplied and Supported by — Fluke International Corporation P.O. Box C9090 Everett, WA 98208 U.S.A. Tel: (206) 358-5500 TLX: 185103 FLUKE UT FAX: 206-356-5118

The following countries are rapresented by: Fluke (Holland) B.V. P.O. Box 2269 5600 CG Eindhoven The Netherlands Tel: (040) 45805, TLX: 51846 FAX: 31-40-457515

Abu Dhabi Mauritania Afghanistan Manorca Albania Mongolla Algeria Nigar Angola Nigeria Senin Qatar Bornholm Rodhos Botswana Chad Sardinia Corsica Saudia Arabia Czechoslavakia Scotland Dubai Senegal Ethiopia Sierra Leone Guinea Somalia lbiza Togo celand Upper Volta Iraq Wales Kuwait Wstn Sahara Liberia Yemen Libya Zaire Luxambourg Zambia Maderia Zimbabwa Mallorca



TECHNICAL SERVICE CENTERS

U.S.A.

CA, Burbank

John Fluke Mfg. Go., air (213) 849-4641

CA, Santa Clara

Jann Fluxe Mig. Co., Inc., (408) 727-8121

CO, Denver

John Fluke Mfg (5) Inc. (303) 695-1000

FL, Oilando

John Fluke Mlg. Co.,Inc. (305) 896-4881

IL, Palatine

John Fluke Mfg Ca . Inc (312) 705-0500

MD, Rockville

John Fluke Mfg. Co.lnc (301) 770-1576

NJ, Paramus

John Fluke Mfg. Co., Inc. (201) 262-9550

TX, Dallas John Flui

John Fluke Mfg Ca., Inc. (214) 869-0311

WA. Seattle

John Fluke Mlg Co., Inc. (206) 356-5560

Other Countries

Argentina, Buenos Aires

Coasin S.A Tel: 552-5248 TLX. (390) 22284

Australia, Brisbane

- Simeasco instruments Pty. Ltd. - Tel. 552-5248

(61.002.0240

Australia, Concord

Elmeasco Instruments Pty Ltd.
Tel: 736-2888

TLX: (790) AA25887

Australia, Ringwood Elmeasco Instruments Pty. Ltd. Tel: 879-2322 TLX: (790) AA36206

Austria, Vienna

Walter Rekirsch Elektronische Gerate GmbH & Co Tel: 253626 TLX: (847) 134759

Belgium, Brussels

Fluke (Belgium) NV/SA Tel: 2164090 TLX: (846) 26312

Brazil, Sao Paulo

Hi-Tek Eletronica Ltda Tel: 421-5477 TLX (1391) 1171913 Canada Calgary

1 tiske Erectronals Carada to: Tei (403) 291-5215 Fax (403) 291-5219

Canada, Mississauga

Fluke Electronics Carvida inc Ter (416) 890-7600 Enx (416) 890-6866

Canada, Montreal

Fruke Flectronics Canada Inn Ter (514) 685-6022 Fax (514) 685-0039

Canada, Ottawa

Fluke Electronics Canada Inc Tel. (613) 723-9453 Fax. (613) 723-9458

Chile, Santiago

Intronica Chile Ltda Ter 232 1886 TLX (332) 346351

China, Peoples Republic of

Instrimpex - Fluke Service Center
Tet: 65-7281

Colombia, Bogota

Sistemas E Instrumentacion, Lída Tel: 232-4532 TLX (396) 45787

Denmark, Ballerup

Tage Olson A/S Tel 658111 TLX (855) 35293

. .

Ecuador, Quito
Proteco Coasin Clal, Etda
Tel: 526759
TEX: (393) 2865

Egypt and Sudan

Electronic Engineering Liaison Old Tel: 2455705 TLX (927) 22782

England, Chertsey, Surrey

Amtest Associates, Lld Tel: 68355

England, Watford, Herts

Fluke (Great Britain) Ltg Tel: 40511 TLX: (851) 934583

Finland

Instrumentarium Elektronikka Tet: 358-0-5281 TLX: (857) 124426

France

M B. Electronique S.A Tel: 1-39-56-81-31 TLX: (842) 695414

Greece, Athens

Hellenic Scientific Representations fel 72*1140 TLX (bb3) 219330

Hong Kong, Wanchai

Schmidt & Co (H.K.) Ltd Tei 8330-222 TLX: (780) 74766 India, Bangalore

Handifron Services Pvt. Ltd. Tel. 363139 Tr. X. (953) 845141

India, Bombay

Hinditron Services Pvt Etd Tel 6300043 Tt X (950) 15-72247

India, New Delhi

Hinddron Services PvI Ltd Tel: 64.33675 TEX: (953) 336458

India, Secunderbad

Hindition Services Pvt. Ltd. Tel 821117 TLX (953) 1556973

Indonesia, Jakarta Pusat

PT Lamda Triguna Tel. 8195365 TLX 46171 LAYARIA

Isiaeł, Tel Aviv

R.D.T. Electronics Engineering Ltd. Ret: 483211 TLX 49221 371452

Italy, Milan

Sistref S.p.A Tel: 6181893 TLX (843) 334643

Japan, Tokyo

John Fluke Mig. Co., Inc. Japan Branch Tel. 434-0181 TLX: (781) 242-4331

Kores, Republic of

Myoung Corp Tel 784-9942 TLX MYOUNG K24283

Malaysia, Selangor

Mecomb Malaysia Sdn. Bhd. Tel: 3-743422 TLX: (784) MA37764

Mexico

Mexicana de Electronica Industrial, S.A. (Mexel) Tel: 5-660-4323 TLX: (383) 1771038

Netherlands, Tilburg

Fluke (Nederland) B.V. Tel: 352455 TLX. (844) 52683

New Zealand, Auckland

Northrop Instruments & Systems Lld. Tel: 501-801 TLX: (791) NZ21570

New Zealand, Wellington

Northrop Instruments & Systems Ltd Tel: 856-658 TLX: (791) NZ3380

Norway, Oslo

Morgenstierne & Co A/S Tel. 356110 TLX (856) 71719 Pakistan, Karachi

International Operations (PAK), Ltd Tel. 221127 | 239052 Tt.X. (952) | 24494

Peru, Lima

Importaciones Y Representaciones Electronicas S A Tel: 288650 TLX (394) 25663

Philippines

Spark Radio & Electronics Corp. Tel. 2-775192 TEX. (722) 27901

Portugal, Lisboa

Equipamentos de Electronica e Científicos, SARL Tel: 2103420 TEX: (832): 15515

Republic of Singapore

Rank O'Connor's (PTE) Limited Tel: 4737944 TLX. (786) RS21023

Republic of South Africa, Bramley

Protes PNI Tel: 11-786-3647 TLX (960) 4-24409

Spain, Madrid

Equipos y Sistemas S A Tel: 458-0150 TLX (831) 42856

Sweden, Vallingby

Kaliber AB Tel 380350 TLX (854) 14077

Switzerland, Zurich

Traco Electronic AG Tel 2010711 TLX (845) 815570

Talwan, Taipei

Schmidt Electronics Corp. Tel: 5013468 TLX. (785) 11111

Thailand, Bangkok

Measuretronix Ltd. Tel: 374-2516 TLX: (788) 82796

Turkey, Islanbul

Erkman Elektronik Aletler Tel: 441546 TLX. (821) 24399

Uruguay, Montevideo

Coasin Uruguaya S.A. Tel: 789015 TLX: (398) UY23010

Venezuela, Caracas Coasin C.A.

Tel: 241-0309 TLX (395) 21027 West Germany, Ismaning/Munich

Fluke (Deutschland) GmbH Tel: 9605240 TLX. (841) 0522472

